



LUND UNIVERSITY

Infinite Horizon Prediction of Post Prandial Breakfast Glucose Excursion

Ståhl, Fredrik; Cescon, Marzia; Johansson, Rolf; Renard, Eric

2009

[Link to publication](#)

Citation for published version (APA):

Ståhl, F., Cescon, M., Johansson, R., & Renard, E. (2009). *Infinite Horizon Prediction of Post Prandial Breakfast Glucose Excursion*. Paper presented at Diabetes Technology Meeting 2009, San Francisco, United States.

Total number of authors:

4

General rights

Unless other specific re-use rights are stated the following general rights apply:

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

Read more about Creative commons licenses: <https://creativecommons.org/licenses/>

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

LUND UNIVERSITY

PO Box 117
221 00 Lund
+46 46-222 00 00



LUND UNIVERSITY

INFINITE HORIZON PREDICTION OF POST PRANDIAL BREAKFAST GLUCOSE EXCURSION

Fredrik Stahl[†] Marzia Cescon Rolf Johansson
Department of Automatic Control, Lund University

Eric Renard

Department of Endocrinology, Montpellier University Hospital

Objective

The objective of the study was to investigate infinite horizon prediction of post prandial blood glucose dynamics after breakfast ingestion using subspace based identification on empirical data.

Data

Five, three days long, patient records (type I diabetes, treated by insulin pump (CSII, $n=3$) or multiple injections (MDI, $n=2$)), collected at the Montpellier Clinical Investigation Center within the European DIAdvisorTM project (www.DIAdvisor.eu), were used. The data records contained frequent fingerprick measurements (average 38 samples/day, HemoCueTM), continuous glucose monitoring (CGM Abbott FreestyleTM) and information on insulin doses and carbohydrate ingestion. The plasma glucose curve was reconstructed from the fingerprick measurements by interpolation. Plasma insulin levels were calculated from the insulin doses using a linear pharmacokinetic model (Wilinska et al. 2005). Using a meal model based upon (Dalla Man et al. 2006) glucose release from the gut following a meal was also assessed.




FIGURE 1: Data record for patient nr 2.

Model Identification

From these data series an individualized, third order state space model of plasma glucose dynamics was identified for each patient using the subspace-based algorithm N4SID in the Matlab Identification Toolbox. Recent previous attempts on model identification and prediction have been made in (Cescon et al. 2009) and (Stahl & Johansson 2009).

The plasma insulin level and gut flux signals, described above, were considered as inputs and the reconstructed plasma glucose curve was used as output. Data from the first night and the following second day breakfast constituted estimation data. The model was cross validated against the first and third breakfasts and compared against the performance of the CGM using Clarke Grid Analysis (pointwise and rate) and RMSE.

Results

Infinite horizon prediction was evaluated for both validation breakfasts for all patients and compared with the performance of the CGM signal in terms of prediction error/measurement error in relation to the reconstructed plasma glucose curve. The infinite horizon predictor outperformed the CGM for all five patients in almost all considered evaluation metrics.




FIGURE 2: HemoCue values, model prediction and CGM in val. data for patient 4.




FIGURE 3: pCGA for model (green) and CGM (blue) for all patients (validation data).




FIGURE 4: pCGA for model (green) and CGM (blue) for all patients (validation data).

Patnr	pCGA(A)		rCGA(A)		RMSE	
	CGM	Model	CGM	Model	CGM	Model
1	74.2	83.9	74.2	81.7	19.2	14
2	96.8	95.2	86.7	93.4	23.6	15.6
3	82.2	100	91.0	93.9	15.0	8.5
4	58.1	98.4	74.8	88.6	45.2	10.1
5	58.9	82.2	79.4	89.9	34.6	24.9
Ave.	74.0	91.9	81.2	89.5	27.5	14.6

TABLE 1: Performance comparison. Amount of points in zone A in pCGA and rCGA in % and RMSE in mg/dl.

Discussion

Statistical analyses show that the difference in pCGA, rCGA and RMSE between the model prediction and CGM is significant. The performance is generally very good, with no points in zone C, D or E for the pCGA. For the rCGA there is one point in zone C. However, this point occurs in the normoglycemic range, and after a short while the model is back on track both in terms of glucose value and direction.

Conclusions

Even though the study is small the results indicate that subspace-based models can be used for prediction of postprandial plasma glucose excursion following breakfasts, with the same accuracy as, or better than, CGM.

Outlook

Further studies will be undertaken on larger data sets to test the methodology in the DIAdvisorTM project within the European FP7-ICT program, IST-216592.

References

- Dalla Man, C., Camilleri, M., Cobelli, C. (2006). A system model of oral glucose absorption: validation on gold standard data. *IEEE TBE*, 53(12), 2472-2477
- Wilinska, M.E., Chassin, L.J., Schaller, H., Schaupp, L., Pieber, T.R., Hovorka, R. (2005). Insulin kinetics in type-1 diabetes: Continuous and bolus delivery of rapid acting insulin. *IEEE TBE*, 52(1), 3-12
- Cescon M., Stahl F., Landin-Olsson M., Johansson R. (2009). Subspace-based Model Identification of Diabetic Blood Glucose Dynamics. *Proceedings of the 15th IFAC Symposium on System Identification*.
- Stahl F., Johansson R. (2009). Diabetes Mellitus Modeling and Short-Term Prediction Based On Blood Glucose Measurements. *Mathematical Biosciences* 217(2), 101-117.

[†] Corresponding author: Fredrik.Stahl@control.lth.se