



LUND UNIVERSITY

Proton Induced X-rays and Gamma Rays for the Analysis of Welding Fumes

Malmqvist, Klas; Johansson, Gerd; Akselsson, Roland

1978

[Link to publication](#)

Citation for published version (APA):

Malmqvist, K., Johansson, G., & Akselsson, R. (1978). *Proton Induced X-rays and Gamma Rays for the Analysis of Welding Fumes*. Abstract from The American Industrial Hygiene Conference, Los Angeles, United States.

Total number of authors:

3

General rights

Unless other specific re-use rights are stated the following general rights apply:

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

Read more about Creative commons licenses: <https://creativecommons.org/licenses/>

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

LUND UNIVERSITY

PO Box 117
221 00 Lund
+46 46-222 00 00

Proton Induced X-rays and Gamma Rays for the Analysis
of Welding Fumes

K.G. Malmqvist¹, G.I. Johansson^{1,2} and K.R. Akselsson^{1,2}

¹Department of Nuclear Physics, University of Lund, Lund, Sweden

²Department of Environmental Health, University of Lund,
Lund, Sweden

Due to current fast progress in material science and welding technology, new, often complex, types of welding aerosols are continually being introduced into work environments. Demands for characterizing welding aerosols from different welding operations and for monitoring welding environments have thus arisen. Proton Induced X-ray Emission analysis, PIXE, constitutes a suitable method for trace element analysis in this context since it provides fast, multielemental analyses at low cost. The mass detection limits are of the order of 0.1-10 ng with an analysing time of 2-5 minutes. Quantitative analysis is possible for elements heavier than phosphorus, the precision and accuracy being about 10 per cent.

When welding with electrodes with low hydrogen coatings, fluorine is a major constituent of the welding aerosol. Detection of this element is important in health applications but it is too light to be analysed with PIXE. However, by including a suitable detection system, we are able to detect simultaneously gamma rays from nuclear reactions for analysing the fluorine content, giving a lower limit of detection of about 100 ng. The low detection limits of this combination of analytical methods permit the design of low-flow (1 l/min) aerosol samplers, suitable for obtaining good particle size- and time-resolution data.

Results from the characterization of welding aerosols will be presented as well as results from measurements at different places.