



LUND UNIVERSITY

Particle Elastic Scattering Analysis of Thin Samples with Protons and He-ions from a 3 MV Electrostatic Accelerator

Bohgard, Mats; Johansson, Eva-Marta

1983

[Link to publication](#)

Citation for published version (APA):

Bohgard, M., & Johansson, E.-M. (1983). *Particle Elastic Scattering Analysis of Thin Samples with Protons and He-ions from a 3 MV Electrostatic Accelerator*. Abstract from Third Conference on Particle Induced X-Ray Emission Analysis and its Applications, Heidelberg, Germany.

Total number of authors:

2

General rights

Unless other specific re-use rights are stated the following general rights apply:

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

Read more about Creative Commons licenses: <https://creativecommons.org/licenses/>

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

LUND UNIVERSITY

PO Box 117
221 00 Lund
+46 46-222 00 00

Abstract from Third Conference on Particle Induced
X-Ray Emission Analysis and its Applications,
Heidelberg, Germany, 1983
C1 - 11:00

PARTICLE ELASTIC SCATTERING ANALYSIS OF THIN SAMPLES WITH PROTONS
AND He-IONS FROM A 3 MV ELECTROSTATIC ACCELERATOR

Mats Bohgard and Eva-Marta Johansson

Department of Nuclear Physics, Lund Institute of Technology, Lund, Sweden

Particle Elastic Scattering Analysis (PESA) is an attractive complement to PIXE for analysis of light elements. It is, however, difficult to find an optimum set of analysis parameters for a wide range of sample types. The elemental resolution is strongly dependent on the mass of the projectile particle, the particle energy, the sample thickness, the scattering angle and the solid angle. The sensitivity of a certain element for a given set of parameters, when using particle energies higher than the Coulomb barrier, can seldom be predicted by theoretical calculations or from published experimental data of scattering yields.

To find optimal conditions for thin aerosol samples and for calibration purpose, we are using polystyrene films, homogeneously doped with different elements to known concentrations. By using cobalt as one of the elements, the simultaneous analysis of the metal by PIXE facilitates the determination of the amount of polystyrene irradiated. Each calibration includes irradiations of polystyrene films with various thickness, which makes it possible to study matrix effects.

The advantages and limitations of PESA with particles from a 3 MV electrostatic accelerator is discussed and the calibration procedure used is described.