

LUND UNIVERSITY

Negative-ion/positive-ion coincidence spectroscopy with a novel spectrometer

Stråhlman, Christian; Sankari, Rami; Kivimäki, Antti; Richter, Robert; Coreno, Marcello; Nyholm, Ralf

Published in: Journal of Physics: Conference Series

DOI: 10.1088/1742-6596/635/11/112123

2015

Link to publication

Citation for published version (APA): Stråhlman, C., Sankari, R., Kivimäki, A., Richter, R., Coreno, M., & Nyholm, R. (2015). Negative-ion/positive-ion coincidence spectroscopy with a novel spectrometer. *Journal of Physics: Conference Series, 635*, Article 112123. https://doi.org/10.1088/1742-6596/635/11/112123

Total number of authors: 6

General rights

Unless other specific re-use rights are stated the following general rights apply: Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

· Users may download and print one copy of any publication from the public portal for the purpose of private study

or research.
You may not further distribute the material or use it for any profit-making activity or commercial gain

· You may freely distribute the URL identifying the publication in the public portal

Read more about Creative commons licenses: https://creativecommons.org/licenses/

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

LUND UNIVERSITY

PO Box 117 221 00 Lund +46 46-222 00 00

Negative-ion/positive-ion coincidence spectroscopy with a novel spectrometer

Christian Stråhlman*¹, Rami Sankari*, Antti Kivimäki[†], Robert Richter[‡], Marcello Coreno[#] and **Ralf Nyholm***

* MAX IV Laboratory, Lund University, Box 118, 221 00 Lund, Sweden [†]CNR-IOM, Laboratorio TASC, 34149 Trieste, Italy [‡] Elettra–Sincrotrone Trieste, Area Science Park, 34149 Trieste, Italy

[#] CNR-ISM, S.S. 14 Km 163.5, Basovizza, I-34149 Trieste, Italy

Synopsis A novel instrument for negative-ion/positive-ion coincidence spectroscopy (NIPICO) is presented. The instrument consists of a double TOF spectrometer with a common extraction region. Electrons are deflected by means of a magnetic field. We demonstrate detections of double (NIPICO) and triple (NIPIPICO) coincidences.

Core-excited states usually decay by emitting electrons or much less probably by emitting soft x-ray photons. The yield of negative ions typically increases at these resonances. Since negative ions only will be created in negativeion/positive-ion pairs, these fragments can be detected in coincidence. Α negativeion/positive-ion coincidence technique (NIPICO) allows us to study how yields of different fragmentation channels vary when the photon energy is tuned to different resonances and associated kinetic energy releases.

We have designed a negative-ion time-of-flight spectrometer to be used together with a positive-ion time-of-flight spectrometer provided by the Gas Phase beamline at the Elettra storage ring [1]. The two spectrometers are mounted facing each other with a common interaction region. Ions and electrons are extracted by means of an electric field. The electrons are deflected by a magnetic field, similar to pioneering experiments by Schermann et al [2], in the negative spectrometer flight tube, reducing efficiently the detection of electron/negative-ion coincidences. The instrument could be operated either using a pulsed electric potential in the interaction region or using a constant field. In the latter case, the identities of the positive and negative ions are deduced from the unique time between the two ion detections.

As a first target for our instrument we chose the SF₆ molecule, known to produce negative ions at resonances close to the S 2p and F 1s edges [3,4]. Figure 1 shows a measurement at two resonances above and below the S 2p ionization edge. They were measured with a constant extraction field and show three main NIPICO channels S⁻/F⁺, F⁻/F⁺ and F^{-}/S^{+} . The pronounced doublet structure of the two former indicates a significant kinetic energy release in the corresponding dissociation channels. A change in branching ratios between the S^{-}/F^{+} and F^{-} $/F^+$ coincidence channels is also seen. We have also

detected several three-particle coincidence channels (NIPIPICO), such as F^{-}/F^{+} and $F^{-}/S^{+}/F^{+}$.



Figure 1. NIPICO spectrum for three resonances close to the S 2p IP. The t-scale indicates the differences between the positive ion TOF and the negative ion TOF. The main coincidence channels are indicated. The sharp peak at t = 0 arise either from crosstalk between the two detectors or fluorescence emission. The second sharp peak at t = -1100 is a computing artefact.

References

[1] O. Plekan et al., Phys. Scr. 78, 058105 (2008)

[2] C. Schremann et al., J. Phys. E: Sci. Instrum., 11, 746 (1978)

[3] S. W. J. Scully et al, J. Phys. B: At. Mol. Opt. Phys. 35, 2703-2711 (2002)

[4] M. N. Piancastelli, et al., J. Chem. Phys. 122, 094312 (2005)

¹ E-mail: Christian.Strahlman@maxlab.lu.se

Content from this work may be used under the terms of the Creative Commons Attribution 3.0 licence. Any further distribution • (cc) of this work must maintain attribution to the author(s) and the title of the work, journal citation and DOI. Published under licence by IOP Publishing Ltd