



LUND UNIVERSITY

Re: Herve Baumert, Andrew Ballaro, Carlos Arroyo, Amir V. Kaisary, Peter F.A. Mulders and Ben C. Knipscheer. The Use of Polymer (Hem-o-lok) Clips for Management of the Renal Hilum During Laparoscopic Nephrectomy. Eur Urol 2006;49:816-9.

Liedberg, Fredrik

Published in:
European Urology

DOI:
[10.1016/j.eururo.2006.06.018](https://doi.org/10.1016/j.eururo.2006.06.018)

2007

[Link to publication](#)

Citation for published version (APA):

Liedberg, F. (2007). Re: Herve Baumert, Andrew Ballaro, Carlos Arroyo, Amir V. Kaisary, Peter F.A. Mulders and Ben C. Knipscheer. The Use of Polymer (Hem-o-lok) Clips for Management of the Renal Hilum During Laparoscopic Nephrectomy. *Eur Urol* 2006;49:816-9. *European Urology*, 51(2), 573-573.
<https://doi.org/10.1016/j.eururo.2006.06.018>

Total number of authors:
1

General rights

Unless other specific re-use rights are stated the following general rights apply:
Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

Read more about Creative commons licenses: <https://creativecommons.org/licenses/>

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

LUND UNIVERSITY

PO Box 117
221 00 Lund
+46 46-222 00 00

This is an author produced version of a paper published in European Urology. This paper has been peer-reviewed but does not include the final publisher proof-corrections or journal pagination.

Citation for the published paper:

Liedberg, F.

"Re: Herve Baumert, Andrew Ballaro, Carlos Arroyo, Amir V. Kaisary, Peter F.A. Mulders and Ben C. Knipscheer. The Use of Polymer (Hem-o-lok) Clips for Management of the Renal Hilum During Laparoscopic Nephrectomy."

European Urology, 2006, Issue: June 30.

[http://dx.doi.org/ 10.1016/j.eururo.2006.06.018](http://dx.doi.org/10.1016/j.eururo.2006.06.018)

Access to the published version may require journal subscription.

Published with permission from: Elsevier

RE: THE USE OF POLYMER (HEM-O-LOK) CLIPS FOR MANAGEMENT OF THE
RENAL HILUM DURING LAPAROSCOPIC NEPHRECTOMY

Baumert H, Ballaro A, Arroyo C, Kaisary A, Mulders P and Knipscheer B. Eur Urol 49
(2006) 816-819.

To the editor: The authors present an elegant DVD describing different techniques for managing the renal hilum during laparoscopic nephrectomy and report on 130 laparoscopic nephrectomies and the use of a 10 mm Hem-o-lok clip. They used the Hem-o-lok clip for the renal artery and on the renal vein after grasping and closing the jaws of a 5-mm laparoscopic Babcock to reduce the diameter of the vein to fit the clips. No significant bleeding from the renal hilum was reported. The arguments for the use of Hem-o-lok clips for management of the renal hilum are lengthening of the renal vein during laparoscopic donor nephrectomy, the cost-effectiveness of the procedure and a more reliable division of the renal vein as compared to the endovascular gastrointestinal stapler are discussed in the article. However, since the first report on the technique by Janetschek and co-workers who ligated the vein prior to clip application in order to diminish the diameter of the vein to fit a 10 mm clip (1), a 15 mm clip has become available (Figure 1). This clip can be applied on almost any renal vein, without using the laparoscopic Babcock as described by the authors in the present series and with minimal handling of the vein, even large right-sided renal veins. The same 15 mm Hem-o-lok clip can be used on the artery, utilizing all six clips in a package in a cost-effective way. I share Baumert and co-workers recommendation to use Hem-o-lok clips for the renal hilum, but suggest 15 mm clips in order to minimize the manipulation of the renal vein prior applying the clips.

Fredrik Liedberg MD, FEBU

Department of Urology

Växjö County Hospital

351 85 Växjö

SWEDEN

Phone: 004647058800

Fax: 0046470588035

f.liedberg@telia.com

1. Janetschek G, Bagheri F, Abdelmaksoud A, Biyani CS, Leeb K, Jeschke S. Ligation of the renal vein during laparoscopic nephrectomy: an effective and reliable method to replace vascular staplers. J Urol 2003;170(4 Pt 1):1295-7.




Figure 1. Ten (below) and fifteen (above) millimeter Hem-o-lok clip in the applicator used for each clip size.