

Time-of-flight laser spectroscopy in biomedical diagnostics

Andersson-Engels, Stefan; Axelsson, Johan; Johansson, Ann; Johansson, Jonas; Svanberg, Sune; Svensson, Tomas

Published in:

Optical Fiber Communication and Optoelectronics Conference, 2007 Asia

DOI:

10.1109/AOE.2007.4410764

2008

Link to publication

Citation for published version (APA):

Andersson-Engels, S., Axelsson, J., Johansson, A., Johansson, J., Svanberg, S., & Svensson, T. (2008). Time-of-flight laser spectroscopy in biomedical diagnostics. In *Optical Fiber Communication and Optoelectronics Conference*, 2007 Asia (pp. 236-238). IEEE - Institute of Electrical and Electronics Engineers Inc.. https://doi.org/10.1109/AOE.2007.4410764

Total number of authors: 6

General rights

Unless other specific re-use rights are stated the following general rights apply:

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.

 • You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

Read more about Creative commons licenses: https://creativecommons.org/licenses/

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Download date: 09. Dec. 2025

Time-of-flight laser spectroscopy in biomedical diagnostics

Stefan Andersson-Engels¹, Johan Axelsson¹, Ann Johansson¹, Jonas Johansson², Sune Svanberg¹ and Tomas Svensson¹

¹Department of Physics, Lund University, PO Box 118, SE-221 00 Lund, Sweden, stefan.andersson-engels@fysik.lth.se

²AstraZeneca Research and Development, Mölndal, Sweden

Optical spectroscopy is becoming a very valuable diagnostic tool in biomedical research. Time-of-flight spectroscopy is a tool providing information regarding scattering and absorption properties of the tissue, valuable for several applications as seen below. In Lund we have in the past developed the technique of supercontinuum generation for spectroscopic investigations of highly scattering media such as tissue[1-3], plant material[4], and pharmaceutical samples[5].

More recently the group in Lund has initiated the development of a unique broadband time-resolved spectroscopy system for turbid media, based on a mode-locked Ti:Sapphire laser pumping a photonic crystal fibre (PCF) and a streak-camera in syncroscan mode.[5,6] The low dispersion of the ultra short laser pulses inside the photonic crystal fibre combined with the small core diameter results in a high peak power of the light in the entire length of the fibre, yielding a high non-linear efficacy resulting in widely spectrally broadened light emission. As a result of this, light pulses with a spectral width spanning from 500 nm to at least 1200 nm were accessible. The streak camera allows recording of time-resolved data with a time resolution of approximately 30 ps. The system is stationary but still relatively flexible and permits both free space as well as fibre coupling of the light onto the sample. This allows different samples and sample geometries to be used. So far direct transmittance (pharmaceutical tablets), fibre coupled transmittance and fibre coupled reflectance (fruits) have been tested. The system has convincingly been used to demonstrate the capability of an analytical instrumentation that could separate absorption and scattering spectra in the evaluation of active substance in pharmaceutical preparations.[7-9] The technique has also been successfully demonstrated for determining the condition of fruits.[10,11]

In contrast to the bulky white light system we have also in parallel been developing a portable diode laser based system for time-resolved measurements.[12] This system is based on the technique called time-correlated single photon counting (TCSPC). This equipment is designed to allow *in vivo* measurement of tissue optical properties of small tissue volumes (less than 1 cm³) at four wavelengths within the tissue optical window (660, 785, 915 and 970 nm). A lot of effort has gone into making the system portable and suitable for a clinical environment. It is now contained within a box of the approximate size of 50*50*30 cm and the light is coupled via optical fibres to and from the patient/sample. This allows the use of sterile fibres to be unpacked and inserted into the

system inside an operation theatre. The system has been used at clinics for unique local characterisation of breast tissue and the prostate gland.[13,14]

Since both systems, produce similar data (although the diode laser based system is confined to only 4 wavelengths) a general evaluation toolbox based on diffusion theory has been developed within the Lund group. The diffusion models allow the calculation of the scattering and absorption coefficients for each wavelength independently. To get the most from these calculated values several approaches have been developed and tested. For example, in the case of the diode based system, the four different absorption coefficients can be used to calculate the concentration of four tissue constituents (oxyhaemoglobin, deoxy-haemoglobin, fat and water).

The multispectral information provided in fluorescence measurements of lesions located at a certain depth in tissue would, in addition to the diagnostic information that there exist a lesion, also provide depth information useful for reconstruction of fluorescent inclusions in tissue[15,16]. This is one possible concept to improving the robustness and accuracy of the fluorescence tomography reconstructions.

One can now foresee a rapidly increasing interest in time-resolved spectroscopic techniques for medical diagnostics, with the development of molecular markers. Most of the genetically specific markers for optical detection will most probably be based on fluorescence. The computer models for tissue fluorescence are not as developed as for remitted light. An accelerated Monte Carlo simulation routine was developed that is approximately 100 times faster than the conventional model used.[17] This model can be used to partly overcome the lack of other appropriate computer models.

Using very similar ideas we are also developing a new technique to perform on-line dosimetric measurements during interstitial photodynamic therapy[18-20]. This development is conducted in close collaboration with SpectraCure AB, a spin-off company from the group. The company has now integrated a system for interstitial photodynamic therapy of prostate cancer. The system is approved for first clinical studies. Our group has developed the technology concept as well as the dosimetry aspects of the treatments to allow treatment feed-back from the measurements.[21,22]. The main challenge is now to further develop the dosimetry aspects of IPDT and to evaluate the procedure by studying treatment outcome.

References

- 1. S. Andersson-Engels, R. Berg, A. Persson, and S. Svanberg, "Multispectral tissue characterization with time-resolved detection of diffusely scattered white light," *Opt. Lett.* **18**, 1697-1699 (1993).
- 2. O. Jarlman, R. Berg, S. Andersson-Engels, S. Svanberg, and H. Pettersson, "Time-resolved white light transillumination for optical imaging," *Acta Radiol.* **38**, 185-189 (1997).
- 3. S. Svanberg, S. Andersson-Engels, R. Cubeddu, E. Förster, M. Grätz, K. Herrlin, G. Hölzer, L. Kiernan, C. af Klinteberg, A. Persson, A. Pifferi, A. Sjögren, and C.-G. Wahlström, "Generation, characterization and medical utilization of laser-produced emission continua," *Laser and Particle Beams* 18, 563-570 (2000).
- 4. J. Johansson, R. Berg, A. Pifferi, S. Svanberg, and L. O. Björn, "Time-resolved studies of light propagation in *Crassula* and *Phaseolus* leaves," *Photochem. Photobiol.* **69**, 242-247 (1999).

- 5. C. Abrahamsson, T. Svensson, S. Svanberg, S. Andersson-Engels, J. Johansson, and S. Folestad, "Time and wavelength resolved spectroscopy of turbid media using light continuum generated in a crystal fiber," *Opt. Express* **12**, 4103-4112 (2004).
- 6. C. Abrahamsson, A. Lowgren, B. Stromdahl, T. Svensson, S. Andersson-Engels, J. Johansson, and S. Folestad, "Scatter correction of transmission near-infrared spectra by photon migration data: Quantitative analysis of solids," *Appl. Spectrosc.* **59**, 1381-1387 (2005).
- 7. C. Abrahamsson, J. Johansson, A. Sparén, and F. Lindgren, "Comparison of Different Variable Selection Methods Conducted on NIR Transmission Measurements on Intact Tablets.," *Chemom. Intell. Lab. Syst.* **69**, 3-12 (2003).
- 8. J. Johansson, S. Folestad, M. Josefson, A. Sparen, C. Abrahamsson, S. Andersson-Engels, and S. Svanberg, "Time-resolved NIR/Vis spectroscopy for analysis of solids: Pharmaceutical tablets," *Appl. Spectrosc.* **56**, 725-731 (2002).
- 9. J. Johansson, S. Folestad, C. Abrahamsson, T. Svensson, and S. Andersson-Engels, "Time-resolved NIR spectroscopy for analysis of solid pharmaceuticals," *NIR News* 17, 4-6 (2006).
- F. Chauchard, S. Roussel, J. M. Roger, V. Bellon-Maurel, C. Abrahamsson, T. Svensson, S. Andersson-Engels, and S. Svanberg, "Least-squares support vector machines modelization for time-resolved spectroscopy," *Appl. Opt.* 44, 7091-7097 (2005).
- F. Chauchard, J. M. Roger, V. Bellon-Maurel, C. Abrahamsson, S. Andersson-Engels, and S. Svanberg, "MADSTRESS: A linear approach for evaluating scattering and absorption coefficients of samples measured using time-resolved spectroscopy in reflection," *Appl. Spectrosc.* 59, 1229-1235 (2005).
- A. Pifferi, A. Torricelli, A. Bassi, P. Taroni, R. Cubeddu, H. Wabnitz, D. Grosenick, M. Möller, R. Macdonald, J. Swartling, T. Svensson, S. Andersson-Engels, R. van Veen, H. J. C. M. Sterenborg, J.-M. Tualle, H. L. Nghiem, S. Avrillier, M. Whelan, and H. Stamm, "Performance assessment of photon migration instruments: the MedPhot protocol," *Appl. Opt.* 44, 2104-2114 (2004).
- 13. T. Svensson, J. Swartling, P. Taroni, A. Torricelli, P. Lindblom, C. Ingvar, and S. Andersson-Engels, "Characterization of normal breast tissue heterogeneity using time-resolved near-infrared spectroscopy," *Phys. Med. Biol.* **50**, 2559-2571 (2005).
- 14. T. Svensson, M. Einarsdóttír, K. Svanberg, and S. Andersson-Engels, "In vivo optical characterization of human prostate tissue using near-infrared time-resolved spectroscopy," *J. Biomedical Optics* **12**, 014022-1-014022-10 (2006).
- 15. J. Svensson and S. Andersson-Engels, "Modeling of spectral changes for depth localization of fluorescent inclusion," *Opt. Express* **13**, 4263-4274 (2005).
- 16. J. Swartling, J. Svensson, D. Bengtsson, K. Terike, and S. Andersson-Engels, "Fluorescence spectra provide information on the depth of fluorescent lesions in tissue," *Appl. Opt.* **44**, 1934-1941 (2005).
- 17. J. Swartling, A. Pifferi, A. M. K. Enejder, and S. Andersson-Engels, "Accelerated Monte Carlo model to simulate fluorescence spectra from layered tissues," *J. Opt. Soc. Am. A* 20, 714-727 (2003).
- 18. A. Johansson, T. Johansson, M. S. Thompson, N. Bendsoe, K. Svanberg, S. Svanberg, and S. Andersson-Engels, "In vivo measurement of parameters of dosimetric importance during interstitial photodynamic therapy of thick skin tumors," *Journal of Biomedical Optics* 11, (2006).
- M. S. Thompson, A. Johansson, T. Johansson, S. Andersson-Engels, S. Svanberg, N. Bendsoe, and K. Svanberg, "Clinical system for interstitial photodynamic therapy with combined on-line dosimetry measurements," *Applied Optics* 44, 4023-4031 (2005).
- M. S. Thompson, S. Andersson-Engels, S. Svanberg, T. Johansson, S. Palsson, N. Bendsoe, A. Derjabo, J. Kapostins, U. Stenram, J. Spigulis, and K. Svanberg, "Photodynamic therapy of nodular basal cell carcinoma with multifiber contact light delivery," *Journal of Environmental Pathology Toxicology and Oncology* 25, 411-424 (2006).
- 21. A. Johansson, N. Bendsoe, K. Svanberg, S. Svanberg, and S. Andersson-Engels, "Influence of treatment-induced changes in tissue absorption on treatment volume during interstitial photodynamic therapy," *Medical Laser Application* **21**, 261-270 (2006).
- A. Johansson, T. Johansson, M. Soto Thompson, N. Bendsoe, K. Svanberg, S. Svanberg, and S. Andersson-Engels, "In vivo measurement of parameters of dosimetric importance during photodynamic therapy of thick skin tumors," *J Biomed Opt.* 11, (2006).