



# LUND UNIVERSITY

## A global proficiency study of Human Papillomavirus genotyping.

Eklund, Carina; Zhou, Tiequn; Dillner, Joakim

*Published in:*  
Journal of Clinical Microbiology

*DOI:*  
[10.1128/JCM.00918-10](https://doi.org/10.1128/JCM.00918-10)

2010

*Document Version:*  
Peer reviewed version (aka post-print)

[Link to publication](#)

*Citation for published version (APA):*  
Eklund, C., Zhou, T., & Dillner, J. (2010). A global proficiency study of Human Papillomavirus genotyping. *Journal of Clinical Microbiology*, 48(11), 4147-4155. <https://doi.org/10.1128/JCM.00918-10>

*Total number of authors:*  
3

### General rights

Unless other specific re-use rights are stated the following general rights apply:  
Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

Read more about Creative commons licenses: <https://creativecommons.org/licenses/>

### Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

LUND UNIVERSITY

PO Box 117  
221 00 Lund  
+46 46-222 00 00

1 A global proficiency study of Human Papillomavirus genotyping

2  
3 Carina Eklund<sup>1</sup>, Tiequn Zhou<sup>2</sup> and Joakim Dillner<sup>1,3\*</sup>

4 for the WHO Human Papillomavirus Laboratory Network

5  
6 WHO HPV LabNet Global Reference Laboratory, Departments of Clinical and Medical  
7 Microbiology, Laboratory Medicine Skåne and Lund University, Malmö, Sweden<sup>1</sup>, Quality, Safety  
8 and Standards Team, Immunizations, Vaccines and Biologicals Department, Family and  
9 Community Health, World Health Organization, Geneva, Switzerland<sup>2</sup> and Departments of  
10 Laboratory Medicine, Medical Epidemiology & Biostatistics, Karolinska Institute, Stockholm,  
11 Sweden<sup>3</sup>

12  
13  
14  
15  
16  
17  
18  
19  
20  
21 \*) Corresponding author:

22 Joakim Dillner, WHO HPV LabNet Global Reference Laboratory, Department of Medical  
23 Microbiology, Lund University, University Hospital MAS, Entrance 78, SE-20502 Malmö,  
24 Sweden. Tel: +46 40 338126. Fax: +46 40 337312. E-mail: [Joakim.dillner@med.lu.se](mailto:Joakim.dillner@med.lu.se)

25 <sup>2</sup>The author is a staff member of the World Health Organization. The author alone is responsible for the  
26 views expressed in this publication and they do not necessarily represent the decisions, policy or views of  
27 the World Health Organization.

28  
29 Running title: WHO Global HPV DNA typing proficiency study  
30  
31

32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56

## ABSTRACT

Internationally comparable quality assurance of Human Papillomavirus (HPV) DNA detection and typing methods is essential for evaluation of HPV vaccines and effective monitoring and implementation of HPV vaccination programs. Therefore, the World Health Organisation (WHO) HPV Laboratory Network (LabNet) designed an international proficiency study. Following announcement at the WHO website, responding laboratories performed HPV typing using one or more of their usual assays on 43 coded samples composed of titration series of purified plasmids of sixteen HPV types (HPV 6, 11, 16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 58, 59, 66 and 68). A detection of at least 50 International Units (IU) of HPV16 or HPV18 DNA and of 500 genome equivalents (GE) of the other 14 HPV types (in samples with single and multiple HPV types) was considered proficient. Fifty-four laboratories worldwide submitted a total of 84 data sets. There were more than 21 HPV genotyping assays used. Commonly used methods were Linear Array, Lineblot, Inno-LiPa, Clinical-Array, type-specific real-time PCR, PCR-Luminex and microarray assays. The major oncogenic HPV types (HPV16 and 18) were detected in 89.7% (70/78) and 92.2% (71/77) of data sets, respectively. HPV types 56, 59 and 68 were the least commonly detected types (in less than 80 % of data sets). Twenty-eight data sets reported multiple false positive results and were considered non-proficient. In conclusion, we found that international proficiency studies, traceable to International Standards, allow a standardised quality assurance of different HPV typing assays and enables a comparison of data generated from different laboratories worldwide.

## INTRODUCTION

57  
58  
59 Human Papillomavirus (HPV) infection is established as the major cause of cervical cancer (2).  
60 Epidemiological studies have classified genital HPV types in high and low-risk HPV types  
61 reflecting their association with invasive cancer (19). The most important high risk types HPV 16  
62 and HPV 18 account for about 70 % of all invasive cervical cancers worldwide. The next most  
63 common HPV types on all continents are HPV 31, 33, 35, 45, 52 and 58 found in approximately 20  
64 % of cervical cancers (19).

65 Accurate and internationally comparable HPV DNA detection and typing methodology is an  
66 essential component in the evaluation of HPV vaccines and in effective implementation and  
67 monitoring of HPV vaccination programs. Genotyping assays used today differ in their  
68 performance with regard to type-specific detection rates (10). As the methodology for quality  
69 assurance and evaluation of assay performance is not standardised, comparisons between different  
70 studies that use different assays is particularly difficult (10).

71 The World Health Organization (WHO) establishes international biological standard materials and  
72 reference reagents for substances of biological origin used in prophylaxis and in therapy or  
73 diagnosis of human diseases ([http://www.who.int/biologicals/reference\\_preparations/en/](http://www.who.int/biologicals/reference_preparations/en/)). At the  
74 WHO meeting held in Geneva, 15-17 August 2005, an expert group recommended the  
75 establishment of a global HPV laboratory network (HPV LabNet), to contribute to improving the  
76 quality of laboratory services for effective surveillance and HPV vaccination impact monitoring.  
77 Major activities within the HPV LabNet include the development of international standard reagents  
78 and standard operating procedures (SOPs), and the development of internationally comparable  
79 quality assurance methods (5, 26).

80 International proficiency panels are already widely used for several microorganisms including  
81 hepatitis A, B and C, herpes simplex virus (HSV) and human immunodeficiency virus (HIV) (15,  
82 18, 24). As there is no natural source of biological material that could be used to generate type-  
83 specific HPV international standards (ISs), the first WHO international collaborative study of  
84 detection of HPV DNA examined the feasibility of using recombinant HPV DNA plasmids as  
85 standards, focusing on HPV 16 and HPV 18 (13). ISs of HPV16 and HPV18 DNA were  
86 established for detection and quantification of HPV 16 and HPV 18 DNA by the WHO Expert  
87 Committee on Biological Standardization in 2008 with assigned potency in International Units  
88 (IU).

89 The international WHO proficiency study described in this report was based on a proficiency panel  
90 composed of purified plasmids containing the genomes of 14 oncogenic HPV types and 2 benign

91 HPV types. As the amount of plasmid DNA was titrated in amounts traceable to the IS, the  
92 proficiency panel allowed an internationally standardised definition of assay sensitivity..  
93 Specificity was defined as absence of incorrect typing. We also evaluated sample pre-processing  
94 with extraction controls of cervical cancer cell lines. The panel was distributed to 61 laboratories  
95 worldwide and analyzed using a range of HPV DNA typing assays in a blinded manner. We report  
96 the results in terms of the ability of participating laboratories to correctly identify HPV types,  
97 grouped by methods performed as well as the analytical sensitivity of detecting the HPV types  
98 included.  
99  
100

## MATERIALS AND METHODS

101  
102  
103  
104 **Source of panel material.** Complete genomes of HPV cloned into plasmid vectors had been  
105 provided to the Lund University by the respective proprietors with a written approval to be used in  
106 this proficiency panel: Dr Ethel-Michele de Villiers (HPV types 6, 11, 16, 18 and 45), Dr Gérard  
107 Orth, (HPV types 33, 39, 66 and 68), Dr Saul Silverstein (HPV type 51), Dr Attila Lörincz (HPV  
108 types 31, 35 and 56), Dr Wayne Lancaster (HPV type 52) and Dr Toshihiko Matsukura (HPV types  
109 58 and 59). The agreements allowed distribution of the plasmids only for the performance of this  
110 WHO proficiency study.

111 The nucleic acid sequences for each of these HPV genomes have been reported previously and are  
112 available in Gene Bank with the following accession numbers; HPV 6 nr X00203; HPV 11 nr  
113 M14119; HPV 16 nr K02718; HPV 18 nr X05015; HPV 31 nr J04353; HPV 33 nr M12732; HPV  
114 35 nr M74117; HPV 39 nr M62849; HPV 45 nr X74479; HPV 51 nr M62877; HPV 52 nr X74481;  
115 HPV 56 nr X74483; HPV 58 nr D90400; HPV 59 nr X77858; HPV 66 nr U31794 and HPV 68 nr  
116 X67161.

117  
118 **Preparation and characterisation of individual panel reagents.** HPV 11 and HPV 58 were  
119 originally cloned in the L1 gene and were therefore re-cloned so that the vector (pGEM4z) is  
120 positioned in the L2 (position 4781) and the E1 (position 1158) gene respectively. For HPV 35 two  
121 clones were included: HPV 35-S contains the entire genes from L1 through E7 including  
122 nucleotides 5012-956, and HPV 35-L including nucleotides 956-5012. The plasmid used for HPV  
123 68 contained only the L1 gene. DNA of each individual HPV genome was generated by the use of  
124 overnight culture of transformed *E. coli* and plasmid purification using Qiagen Midi-prep kit.  
125 Optical density determinations were made at 260 nm and 280 nm to estimate purity of the  
126 preparation. Size and purity of the plasmids were analysed using agarose gel electrophoresis. The  
127 double stranded DNA concentration was established using fluorimetric measurements by picogreen  
128 quantitations (PicoGreen dsDNA Quantitation Reagent; Molecular Probes, Inc, Eugene, Oreg). The  
129 purified plasmid bulk of HPV 16 and HPV 18 were tested in ten-fold serial dilutions in parallel  
130 with International Standards for HPV 16 (06/202) and HPV 18 (06/206) distributed by NIBSC  
131 (Hertfordshire, UK) using a PCR Luminex assay to establish the amounts in terms of International  
132 Units, by traceability of the amount of plasmids in the panel to the IS (16).

133  
134 **Panel composition and production.** Purified plasmids containing cloned genomic DNAs for HPV  
135 types 6, 11, 16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 58, 59, 66 and 68 were diluted to a stock

136 concentration of  $10^8$  genome equivalents/ $\mu\text{l}$  in TE buffer (10mM Tris, 1 mM EDTA pH 8.0) to be  
137 used for preparation of 43 samples. Human placenta DNA (Sigma-Aldrich 7011) at a concentration  
138 of 10 ng/ $\mu\text{l}$  was added to the TE buffer to mimic a molecular matrix background that would  
139 typically be present in biological samples. Table 1 summarizes the composition of the panel. The  
140 different amounts of plasmid (5-500 GE or IU) were chosen to reflect the lower spectrum of  
141 amount of virus that would typically be present in clinical samples. E.g., a study of virus quantities  
142 present in cervical samples from healthy HPV-positive women found an average of 18000 GE of  
143 HPV16/100ng input DNA (range  $<300 - 14000000$  GE/100ng input DNA) (20). The 43 different  
144 panel samples were prepared by dilution of HPV recombinant DNA plasmid stock solution in TE  
145 buffer in the background of human placental DNA. Briefly, the HPV DNA plasmids were diluted  
146 100-fold in TE-placenta buffer to  $10^4$  genome equivalents (GE)/ $\mu\text{l}$ , further 10 fold dilutions were  
147 made to a final concentration of 1 IU/ $\mu\text{l}$  of HPV 16 and HPV 18, for the other HPV types included  
148 10 GE/ $\mu\text{l}$  was the final dilution. To ensure high quality of the panel two HPV types were diluted  
149 each day with an interval of at least 4 hours in between. The samples containing multiple types  
150 were prepared from dilutions of  $10^3$  genome equivalents/ $\mu\text{l}$ . After production of each of the 43  
151 reference samples, the preparation was dispensed in 100  $\mu\text{l}$  volumes in 1.5 ml siliconized vials. The  
152 vials were labelled as WHO HPV DNA 2008 and randomly assigned numbers from 1 through 43.  
153 The panels were stored at  $+4^\circ\text{C}$  before shipment to participating laboratories. Participants were  
154 instructed to perform HPV typing according to their standard methods using their standard sample  
155 input volume.

156 Two different cell lines were used as controls of the extraction process in participating laboratories.  
157 The HPV-negative epithelial cell line C33A derived from human cervical carcinoma and the  
158 HPV16-positive epithelial cell line SiHa, derived from a squamous cell carcinoma was purchased  
159 from the American Type Culture Collection and cultured in Dulbecco's modified Eagle medium  
160 (Gibco 11960). The cells were diluted in PreserveCyt<sup>TM</sup> (Cytoc 0234004) to a concentration of 400  
161 cells/ $\mu\text{l}$ , 100  $\mu\text{l}$  of each preparation was dispensed in 1.5 ml vials and labelled WHO HPV DNA A  
162 and B.

163 Before distribution of the WHO HPV DNA proficiency panel, the samples were tested (blinded) at  
164 the WHO HPV LabNet Global Reference Laboratory (GRL) in Sweden and two other laboratories,  
165 namely the German Cancer Research Center (DKFZ) in Heidelberg, Germany (Dr. Michael  
166 Pawlita) and the WHO HPV LabNet GRL at Centers for Disease Control and Prevention (CDC) in  
167 the United States (Dr. Elizabeth Unger).

168

169 **Technologies used for initial characterization of the panel. (i) GRL Sweden.** Three  
170 independent experiments testing each sample in duplicate were performed. Five microliters of  
171 panel sample DNAs was used for MGP PCR as previously described (21). Ten  $\mu$ l PCR products  
172 were analysed by multiplex genotyping using a Luminex based assay as described earlier (16, 17).  
173 HPV types 6, 11, 16, 18, 26, 30, 31, 33, 35, 39, 40, 42, 43, 45, 51, 52, 53, 54, 56, 58, 59, 66, 67,  
174 68a, 69, 70, 73, 74, 82, 86, 89, 90 and 91 were distinguished. Appropriate negative and positive  
175 controls were used to monitor the performance of the method. DNA from the extraction control A  
176 and B was extracted using QIAamp DNA Mini and Blood kit (Qiagen) according to the  
177 manufacturer's instructions.

178 **(ii) DKFZ.** A 10  $\mu$ l DNA sample was amplified by the broad-spectrum GP5+ / 6+ primers as  
179 previously described (17). The PCR products were analysed using bead based multiplex  
180 genotyping as described (16). HPV types 6, 11, 16, 18, 26, 30, 31, 33, 35, 39, 42, 43, 44, 45, 51,  
181 52, 53, 56, 58, 59, 66, 67, 68a, 68b (Me 180), 69, 70, 73 and 82 were distinguished. All samples  
182 were tested for human DNA with PCR primers amplifying part of the  $\beta$ -globin gene and a bead  
183 coupled  $\beta$ -globin specific probe used in the genotyping assay.

184 **(iii) GRL CDC.** Ten microliters of sample DNAs was used in the 100  $\mu$ l PCR otherwise following  
185 the manufacturers protocol for Roche Linear Array which is designed to detect 37 individual HPV  
186 types, 6, 11, 16, 18, 26, 31, 33, 35, 39, 40, 42, 45, 51, 52, 53, 54, 55, 56, 58, 59, 61, 62, 64, 66, 67,  
187 68, 69, 70, 71, 72, 73, 81, 82, 83, 84, 89 and IS39. As the probe for detecting HPV 52 cross-  
188 hybridizes to types 33, 35 and 58, the presence of HPV 52 in samples with one or more of these  
189 three other types was tested with an HPV 52-specific real-time PCR assay.

190  
191 **Organization of the study.** Participants to the study were recruited by advertisement at the WHO  
192 website. The panels were distributed from the WHO HPV LabNet GRL in Sweden at ambient  
193 temperature to 61 laboratories worldwide, by WHO region: America Region 16 laboratories, Africa  
194 Region 1 laboratory, Eastern Mediterranean Region 1 laboratory, European Region 28 laboratories,  
195 South East Asia Region 2 laboratories and Western Pacific Region 13 laboratories. The package  
196 also included a letter of instruction as well as a form for reporting the results of the testing of the  
197 panel as well as technical information on the procedures to be performed. Laboratories were asked  
198 to submit the results of the tests performed to the WHO GRL in Sweden within 4 weeks of receipt  
199 of specimens. The agreement included assigning the right to publish the data to the WHO, but it  
200 was agreed that only coded results from all laboratories will be presented, grouped by methods  
201 performed.



202  
203  
204  
205  
206  
207  
208  
209  
210  
211  
212  
213  
214  
215  
216  
217  
218  
219  
220  
221  
222  
223  
224

All results submitted to the WHO HPV LabNet GRL Sweden were coded and analyzed anonymously. Data sets generated were designated numerically from 1 through 84. Individual results of the proficiency study were disclosed only to the participating laboratory that generated the data.

**HPV technologies used by study participants.** The different HPV typing methods that were used to generate results for the WHO LabNet proficiency study to detect HPV DNA (1, 3, 6-9, 17, 21, 22, 25) are summarized in Table 2.

**Data analysis.** Criteria used for considering a data set as proficient were the following: (i) detection of at least 50 IU per 5 µl of HPV 16 and HPV 18, both in single and multiple HPV infection; (ii) detection of at least 500 GE per 5 µl of the other HPV types included, both in single and multiple infection; (iii) at most one false positive result. These criteria were arrived at by a consensus opinion of international experts participating in an international WHO workshop in Geneva, 2008 (5) and was based on a consideration of which performance requirements were required and realistic. A higher requirement for HPV16 and 18 was considered essential, because of the pivotal role of these HPV types in causing cervical cervical cancer.

Four data sets reporting results only as “high” or “low risk” HPV were not included in the overall performance analyses (one data set that used the Roche Amplicor assay, one data set that used the Seeplex HPV 4 ACE assay and two data sets used in-house PCR with agarose gel analyses).

## RESULTS

225  
226  
227

228 **Validation of the HPV proficiency panel.** The results from the initial panel validation at the 2  
229 GRLs and at DKFZ included qualitative characterization of HPV and human genomic DNAs. Two  
230 of these laboratories used Luminex based assays with modified GP 5+ / 6+ primers, the third  
231 laboratory used Linear Array, which is based on PGMY primers for the analyses. No false positive  
232 HPV type was detected in the samples in any of the reference laboratories. HPV 31 was not  
233 detected in the lowest concentration, when present together with other plasmids, in both  
234 laboratories that used GP-based assays. HPV 18 was not detected in the lowest concentration in the  
235 laboratory using Linear Array. All other HPV types were detected at the lowest concentration  
236 included in the panel, except HPV 39 and HPV 68 that could not be detected using Linear Array.  
237 The HPV 39 plasmid used in the panel cannot be detected by systems using PGMY primers as it  
238 was cloned into the vector at the binding site of one of the PGMY primers. Linear Array and other  
239 PGMY based assays are designed to detect HPV68 subtype B and can not detect the HPV68  
240 prototype virus because of several mismatches.

241 All 3 reference laboratories detected HPV 16 DNA in the DNA extraction control containing SiHa  
242 cells and had negative results in the negative control for DNA extraction (C33A cells).

243 The results from the reference laboratory evaluation advised that the panel performed as expected  
244 and the panel was then distributed to participating laboratories worldwide.

245

246 **Panel distribution and response.** Fifty-four of 61 participating laboratories, including the three  
247 laboratories who did the panel validation, submitted 84 data sets according to the timeline (Table  
248 2). Two laboratories that responded after the deadline are not included in this report. Four data sets  
249 were generated using assays that did not discriminate specific HPV types and were therefore not  
250 included in the overall type-specific analyses presented here.

251 Some participating laboratories did not perform tests for typing of all HPV types included in the  
252 proficiency panel. Therefore, the denominator for the number of test results included in the  
253 analyses varies for the different HPV types. In 37 data sets, the results had been obtained using  
254 commercially available tests. The most commonly used assay was Linear Array (Roche) that was  
255 used to generate 15 data sets. Other widely used assays were CLART HPV 2 (Genomica), Inno-  
256 LiPA (Innogenetics), PGMY-LINEBLOT, in-house type specific PCR, Luminex and Microarray  
257 based assays (Table 2). Participating laboratories included both public health laboratories, research  
258 laboratories, diagnostic kit manufacturers and vaccine companies. The annual number of samples

259 analysed for HPV per laboratory varied from 100 to 100 000 per year with approximately 40 % of  
260 the laboratories performing <2000 HPV typing tests per year and around 40 % between 2000 and  
261 10 000 HPV typings per year.

262  
263 **Performance of HPV assays and participating laboratories.** Participating laboratories were  
264 requested to perform testing using their standard protocols. Accordingly, the input volume of the  
265 DNA panel varied between 2 µl and 50 µl between laboratories. Data is presented by lowest  
266 category of concentration (5, 50 or 500 GE or IU) proven to be detectable. E.g., a lab using a 2 ul  
267 input instead of 5ul input that does detect 2 GE is considered to be able to detect 5 GE. The sample  
268 containing 100 IU HPV 16/µl was the sample that most datasets, 94.9 %, identified correctly  
269 (Table 1). Single HPV types in 100 GE/µl were correctly identified, without false positive types  
270 detected, in an average of 84 % of the data sets. HPV 56 and 59 were correctly identified by less  
271 than 80 % of the datasets HPV 68 was correctly identified only by 37.9 % of laboratories. In the  
272 samples containing multiple HPV types, between 50 % and 73 % of the datasets could correctly  
273 identify the types. The negative control sample containing only human genomic DNA was  
274 correctly identified as negative by 74 of 80 datasets.

275 The proficiency of detecting HPV types (restricted to data sets testing for more than 12 HPV types)  
276 is shown in Table 2. Nineteen data sets were 100 % proficient (detecting at least 50 IU of HPV 16  
277 and HPV 18 in 5µl and 500 GE in 5µl of the other HPV types tested for (also when present  
278 together with other HPV types), without having more than one false positive result. As the Linear  
279 Array assays used a large (50 µl) input volume, the Linear Array system did not test for presence  
280 of amounts below 50 IU of HPV 16 and HPV 18 in 5µl and 500 genome equivalents in 5µl of the  
281 other HPV types. Two different Microarray assays were the commercial tests that had the highest  
282 number of proficient results (100%). Several in-house assays based on type-specific PCR and on  
283 general-primer PCR-Luminex were also 100 % proficient.

284 The non-commercial PGMY-LineBlot assay was transferred to all WHO HPV LabNet members in  
285 2008 as an effort to build up testing capacity and evaluate the ease of technology transfer of this  
286 assay. The PGMY-LineBlot assay was used by seven members of the WHO HPV LabNet but with  
287 100 % proficiency in only one laboratory. Only one laboratory (the originator) had been routinely  
288 using this assay before and the other laboratories had recently set up the assay according to  
289 instructions. Indeed, when a subsequent, similar proficiency panel was sent to the WHO HPV  
290 LabNet members, two additional laboratories using PGMY-LineBlot were 100% proficient and  
291 one additional lab was 88% proficient (data not shown).

292 To be considered as proficient in this study no more than one false positive sample per data set was  
293 acceptable. The number of false positive HPV types detected per data set is shown in Table 3.  
294 Thirty-four of the 80 data sets did not have any false positive results, whereas 12 data sets reported  
295 more than 3 false positive results. Among these, 3 datasets reported false positive HPV types in  
296 more than 15 samples. Data sets generated by the commercial tests CLART and InnoLiPA reported  
297 more than one false positive sample in 4 out of 6 datasets. Several in-house assays as well as some  
298 commercial assays that were performed by only one or only a few laboratories reported no false  
299 positive results at all.

300 The lowest genome equivalent or IU of each HPV type included in the panel that was detected in  
301 both single and multiple infections by different assays are shown in Table 4. HPV 16 and HPV 18  
302 were the types detected at lowest concentration in most data sets. Only 1 and 3 datasets,  
303 respectively, could not detect the highest concentration of HPV16 and 18. By contrast, for HPV 52,  
304 HPV 59 and HPV 56 there were 25, 19 and 17 data sets, respectively, that could not detect these  
305 viruses in the highest concentration (Table 4).

306  
307  
308  
309

## DISCUSSION

310  
311  
312 We report on the development of an internationally comparable quality assurance methodology  
313 that traceable to ISs. Accurate and internationally comparable HPV DNA detection and typing  
314 methodology is an essential component in the evaluation of HPV vaccines and in effective  
315 implementation and monitoring of HPV vaccination programmes. Standardised methodology for  
316 evaluation of laboratory performance is a fundamental basis to enable any comparison of the  
317 methodologies used in laboratories worldwide. The major tools for achieving progress towards  
318 this goal are developing international biological standards and preparing and validating proficiency  
319 panels to qualify methods. The current study has established that such international proficiency  
320 panels with unitage traceable to ISs can be used in global studies. We have also demonstrated that  
321 such studies provide a unique overview of the status of the HPV detection and typing methodology  
322 that are being used globally and how well they perform in different laboratories.  
323  
324 Overall, it can be said that a majority of laboratories in this study had a good performance of their  
325 HPV DNA typing tests. However, some limitations were revealed.  
326 There was a clear tendency towards systematically different limits of sensitivity for different HPV  
327 types. E.g., HPV 16 and HPV 18 were the types detected at the lowest amount in most data sets  
328 (only 1 and 3 datasets, respectively, could not detect 500 IU / 5 µl) whereas HPV 52, HPV 59 and  
329 HPV 56 could not be detected in the 500 GE / 5 µl amount by 25, 19 and 18 data sets respectively.  
330 Thus, many surveys of circulating HPV types might systematically underestimate the prevalence of  
331 HPV 52, 56 and 59 compared to HPV 16 and 18.  
332 There was also a tendency for lower sensitivity of tests when multiple HPV types were present. In  
333 the samples containing multiple HPV types, between 50 % and 73 % of the data sets could  
334 correctly identify the types present, but in samples with only 1 HPV type present an average of 84  
335 % of HPV types could be identified without false positive results. This tendency will cause a  
336 systematic underestimation of the prevalence of multiple infections and will introduce a systematic  
337 detection bias in epidemiological studies with detectability being dependent on determinants of  
338 HPV acquisition (e.g., a given HPV type will be more difficult to detect in high risk groups,  
339 because of higher likelihood of other HPV infections).  
340 There was a surprisingly high amount of false positive results reported, with only 34/80 datasets  
341 being 100% specific. The proficiency panel contained only 1 entirely HPV-negative sample. The  
342 present study was designed to primarily evaluate HPV typing (rather than mere HPV detection)  
343 and we considered that specificity should in this context be measured primarily as absence of

344 detection of a specific HPV type also when other HPV types are present. Thus, for each HPV type  
345 evaluated there are at least 39 negative samples included in the panel and 1 false positive result  
346 thus equals >97% specificity. There was only 1 indication of a systematic mistyping (some Linear  
347 Array-based data sets reporting HPV56-containing samples as positive for HPV66), but otherwise  
348 there was no single sample that had systematic false positivity for the same type in several  
349 laboratories. These very common false positives are therefore neither associated with the panel nor  
350 with the assays used, but rather appear to result from laboratory environment and performance.  
351 Considering the deleterious consequences that a false positive result may have, it appears that a  
352 substantial effort towards increased specificity of testing is warranted.

353  
354 On the other hand, there were some needs for improvement of the proficiency panel itself that were  
355 identified by this study. The HPV 39 plasmid used in the panel was cloned into the vector at the  
356 binding site of one of the most commonly used PCR primers (PGMY). All assays using the PGMY  
357 primer system, including Linear Array and CLART, could not detect the HPV 39 plasmid in the  
358 panel. As this was because of the way the plasmid was constructed, all these data sets were  
359 considered as not having been evaluated for HPV39 in this study.

360 The plasmid used to test for HPV 68a was not full-length, but contained only the L1 gene. We  
361 noted that Linear Array and all other PGMY-based assays that are indeed directed against L1 could  
362 not detect the HPV68a plasmid. Comparison of the sequences of HPV68a and HPV68b (ME180  
363 isolate) showed significant differences in the sequence corresponding to the PGMY primer binding  
364 site. As the sequence of HPV68b was published before the sequence of HPV68a, it appears that  
365 these systems are designed to only detect HPV68b (11, 14). All data sets reporting usage of  
366 primers directed to genes other than L1 or that used the PGMY primers were considered as not  
367 testing for HPV 68 in this study. Accordingly only 29 data sets could be analysed for detection of  
368 HPV 68a and only 11 of the 29 laboratories (38 %) could detect HPV 68a. For the next WHO  
369 HPV LabNet proficiency panel, HPV39 will be re-cloned to change the cloning site and full-length  
370 genomes of both HPV68a and HPV68b will also be included.

371  
372 The Linear Array can not exclude HPV 52 when the sample is positive for HPV 33, HPV 35 or  
373 HPV 58. Some laboratories have developed a type-specific PCR for HPV 52 to test HPV 33, 35  
374 and 58-positive samples, whereas some laboratories (4/15) scored all sample with multiple  
375 infections containing HPV 52 as negative for HPV 52 (4, 23). This resulted in that they are  
376 regarded as not proficient for HPV 52 in this study. Four data sets generated using Linear Array  
377 were considered as not proficient since they reported 2 or even 3 false positive results. HPV 66 was

378 detected as false positive in 7 of in total 15 false positive results submitted in the 15 data sets using  
379 Linear Array, 6 of these samples contained 500 GE of HPV 56 that was correctly identified. The  
380 detection of HPV 66 in these samples was not reported by any other assay, indicating that the false  
381 detection of HPV 66 in HPV 56-positive samples is a problem that is commonly seen with the  
382 Linear Array assay.

383 For two commercial tests (InnoLiPA and CLART), 4 out of 6 data sets were not proficient because  
384 of too many false positives. InnoLiPa could not identify HPV 52 in 5 of 6 data sets. On the other  
385 hand, HPV 52 was reported in 9 samples where it was not present. The number of false positive  
386 samples reported by InnoLiPA was between 3 and 5 for the 4 laboratories that were not proficient.  
387 Three laboratories using CLART reported 7, 17 and 21 false positive results respectively, some  
388 with more than 3 false positives in each sample. Four laboratories using CLART could not detect  
389 HPV 56 and 45 in samples with multiple types. There was no consistent false positivity for any  
390 specific sample for these two assays. The false positivities for these assays appeared to be  
391 randomly distributed among the samples and were always different for the different laboratories,  
392 indicating that the problem is not related to the assay kit itself. Indeed, there were examples of  
393 several laboratories that had completely proficient results using these assays.

394  
395 A major conclusion of the present study is that differences in performance were much larger  
396 between laboratories than between different types of assays. Proficiency panel testing is  
397 particularly useful to stimulate a learning process of improved performance in laboratories. Once  
398 regular feed-back on proficiency testing results is implemented, improvement of performance  
399 usually follows rapidly. An example of this was the results of the PGMY-LineBLOT assay that  
400 was recently set up in the HPV LabNet. Several laboratories who were using this assay for the first  
401 time had suboptimal results, but became proficient in a subsequent proficiency testing performed  
402 when there had been more time for practise.

403  
404 The 2 samples that evaluated the DNA extraction step before the HPV testing and typing had a  
405 surprisingly low proportion of correct results. The sample containing 2000 cells of the cervical  
406 cancer cell line SiHa with about 1 copy of HPV16 per cell (i.e. total 2000 IU of HPV16/5ul) was  
407 detected only in about a third of the datasets. Also, a large number of datasets (six) reported false  
408 positive results in the sample containing an HPV-negative human cell line. This indicates that low  
409 yield in the DNA extraction step, potentially reducing sensitivity, as well as contamination in the  
410 DNA extraction step may be significant problems in the field of HPV DNA testing. Future

411 proficiency panels will contain a larger set of samples designed to specifically evaluate the DNA  
412 extraction step before the actual HPV testing and typing.  
413 There are additional steps in the laboratory detection process that are not evaluated by the present  
414 strategy, notably sampling technique, handling and storage, natural variability of circulating virus  
415 strains, PCR inhibiting substances and naturally occurring genome modifications (e.g. integration  
416 and rearrangement). The HPV LabNet has chosen to perform quality control for these aspects of  
417 testing by launching a confirmatory testing scheme, where part of the clinical samples being tested  
418 are annually submitted for retesting to a higher level reference laboratory (5). The alternative  
419 strategy to include clinical samples in proficiency testing schemes was not chosen, because of the  
420 need to have exactly reproducible panels with defined content that can be used by hundreds of  
421 laboratories over many years and since confirmatory testing schemes were considered to better  
422 reflect the actual testing being done.

423  
424 It should be emphasised that the current proficiency panel study was designed to evaluate the  
425 performance of HPV testing and typing tests used in HPV vaccinology and HPV surveillance, but  
426 not for evaluation of HPV tests used in cervical cancer screening (12). The demands on  
427 performance of HPV typing assays vary depending on the purpose of the testing. In vaccinology, a  
428 high sensitivity is needed for clinical vaccine trials as failure to detect prevalent infections at entry  
429 may result in apparent vaccine failures. By contrast, the clinical HPV-associated diseases, such as  
430 high grade CIN, typically contain larger amounts of virus and cervical screening programs using  
431 HPV testing do not have as high demands on sensitivity (12). Guidelines for evaluations of such  
432 tests have recently been published (12).

433  
434 In conclusion, we find that global HPV DNA proficiency studies are both feasible and informative.  
435 The launch of an internationally standardised methodology to analyse the specificity and sensitivity  
436 for different HPV typing assays (as well as the performance of participating laboratories) to  
437 correctly identify the 16 HPV types which are the most important in HPV surveillance and  
438 vaccinology is likely to greatly enhance quality and comparability of studies in these fields.

439  
440 **ACKNOWLEDGMENTS**  
441 Members of the WHO HPV LabNet are as follows: A.C. Bharti, Division of Molecular Oncology,  
442 Institute of Cytology and Preventive Oncology, Nodia, India; J. Dillner, Department of Medical  
443 Microbiology, Lund University, Malmo, Sweden; E. Ennaifer-Jerbi, Tunis Pasteur Institute, Tunis,  
444 Tunisia; S. Garland, Department of Microbiology and Infectious Disease, Royal Women's



445 hospital, Carlton, Australia; I Kukimoto, Centre for Pathogen Genomics, National Institute of  
446 Infectious Diseases, Tokyo, Japan; A.M. Picconi, National Institute of Infectious Diseases, Buenos  
447 Aires, Argentina; R. Sahli, Institut de Microbiologie, CHUV, Lausanne, Switzerland; S. Sukvirach,  
448 National Cancer Institute, Bangkok, Thailand; E.R. Unger, Centres for Disease Control and  
449 Prevention, Atlanta, GA, USA; and A.L Williamson, Institute of Infectious Disease and Molecular  
450 Medicine, University of Cape town, Cape Town, South Africa.

451 We thank Michael Pawlita of the German Cancer Research Center (DKFZ) for acting as external  
452 validation laboratory in evaluation of the proficiency panel. We also thank all the 61 participants  
453 worldwide that participated in the proficiency study.

454 This work was supported by the WHO via a project funded by Bill & Melinda Gates foundation.

- 457 1. **Bernard, H. U., S. Y. Chan, M. M. Manos, C. K. Ong, L. L. Villa, H. Delius, C. L.**  
458 **Peyton, H. M. Bauer, and C. M. Wheeler.** 1994. Identification and assessment of known  
459 and novel human papillomaviruses by polymerase chain reaction amplification, restriction  
460 fragment length polymorphisms, nucleotide sequence, and phylogenetic algorithms. *J Infect*  
461 *Dis* **170**:1077-85.
- 462 2. **Bosch, F. X., A. Lorincz, N. Munoz, C. J. Meijer, and K. V. Shah.** 2002. The causal  
463 relation between human papillomavirus and cervical cancer. *J Clin Pathol* **55**:244-65.
- 464 3. **Chouhy, D., L. B. Gil, A. L. Nocito, D. Wojdyla, L. Ornella, J. Cittadini, D. Gardiol,**  
465 **and A. A. Giri.** 2006. Development and evaluation of a colorimetric PCR system for the  
466 detection and typing of human papillomaviruses. *Int J Mol Med* **18**:995-1003.
- 467 4. **Coutlee, F., D. Rouleau, G. Ghattas, C. Hankins, S. Vezina, P. Cote, J. Macleod, A. de**  
468 **Pokomandy, D. Money, S. Walmsley, H. Voyer, P. Brassard, and E. Franco.** 2007.  
469 Confirmatory real-time PCR assay for human papillomavirus (HPV) type 52 infection in  
470 anogenital specimens screened for HPV infection with the linear array HPV genotyping  
471 test. *J Clin Microbiol* **45**:3821-3.
- 472 5. **Ferguson, M., D. E. Wilkinson, and T. Zhou.** 2009. WHO meeting on the standardization  
473 of HPV assays and the role of the WHO HPV Laboratory Network in supporting vaccine  
474 introduction held on 24-25 January 2008, Geneva, Switzerland. *Vaccine* **27**:337-47.
- 475 6. **Gheit, T., S. Landi, F. Gemignani, P. J. Snijders, S. Vaccarella, S. Franceschi, F.**  
476 **Canzian, and M. Tommasino.** 2006. Development of a sensitive and specific assay  
477 combining multiplex PCR and DNA microarray primer extension to detect high-risk  
478 mucosal human papillomavirus types. *J Clin Microbiol* **44**:2025-31.
- 479 7. **Gravitt, P. E., C. L. Peyton, R. J. Apple, and C. M. Wheeler.** 1998. Genotyping of 27  
480 human papillomavirus types by using L1 consensus PCR products by a single-  
481 hybridization, reverse line blot detection method. *J Clin Microbiol* **36**:3020-7.
- 482 8. **Jacobs, M. V., A. M. de Roda Husman, A. J. van den Brule, P. J. Snijders, C. J.**  
483 **Meijer, and J. M. Walboomers.** 1995. Group-specific differentiation between high- and  
484 low-risk human papillomavirus genotypes by general primer-mediated PCR and two  
485 cocktails of oligonucleotide probes. *J Clin Microbiol* **33**:901-5.
- 486 9. **Kleter, B., L. J. van Doorn, L. Schrauwen, A. Molijn, S. Sastrowijoto, J. ter Schegget,**  
487 **J. Lindeman, B. ter Harmsel, M. Burger, and W. Quint.** 1999. Development and clinical  
488 evaluation of a highly sensitive PCR-reverse hybridization line probe assay for detection  
489 and identification of anogenital human papillomavirus. *J Clin Microbiol* **37**:2508-17.
- 490 10. **Klug, S. J., A. Molijn, B. Schopp, B. Holz, A. Iftner, W. Quint, J. F. S. P, K. U. Petry,**  
491 **S. Kruger Kjaer, C. Munk, and T. Iftner.** 2008. Comparison of the performance of  
492 different HPV genotyping methods for detecting genital HPV types. *J Med Virol* **80**:1264-  
493 74.
- 494 11. **Longuet, M., S. Beaudenon, and G. Orth.** 1996. Two novel genital human papillomavirus  
495 (HPV) types, HPV68 and HPV70, related to the potentially oncogenic HPV39. *J Clin*  
496 *Microbiol* **34**:738-44.
- 497 12. **Meijer, C. J., J. Berkhof, P. E. Castle, A. T. Hesselink, E. L. Franco, G. Ronco, M.**  
498 **Arbyn, F. X. Bosch, J. Cuzick, J. Dillner, D. A. Heideman, and P. J. Snijders.** 2009.  
499 Guidelines for human papillomavirus DNA test requirements for primary cervical cancer  
500 screening in women 30 years and older. *Int J Cancer* **124**:516-20.
- 501 13. **Quint, W. G., S. R. Pagliusi, N. Lelie, E. M. de Villiers, and C. M. Wheeler.** 2006.  
502 Results of the first World Health Organization international collaborative study of detection  
503 of human papillomavirus DNA. *J Clin Microbiol* **44**:571-9.

- 504 14. **Reuter, S., H. Delius, T. Kahn, B. Hofmann, H. zur Hausen, and E. Schwarz.** 1991.  
505 Characterization of a novel human papillomavirus DNA in the cervical carcinoma cell line  
506 ME180. *J Virol* **65**:5564-8.
- 507 15. **Schloss, L., A. M. van Loon, P. Cinque, G. Cleator, J. M. Echevarria, K. I. Falk, P.**  
508 **Klapper, J. Schirm, B. F. Vestergaard, H. Niesters, T. Popow-Kraupp, W. Quint, and**  
509 **A. Linde.** 2003. An international external quality assessment of nucleic acid amplification  
510 of herpes simplex virus. *J Clin Virol* **28**:175-85.
- 511 16. **Schmitt, M., I. G. Bravo, P. J. Snijders, L. Gissmann, M. Pawlita, and T. Waterboer.**  
512 2006. Bead-based multiplex genotyping of human papillomaviruses. *J Clin Microbiol*  
513 **44**:504-12.
- 514 17. **Schmitt, M., B. Dondog, T. Waterboer, and M. Pawlita.** 2008. Homogeneous  
515 amplification of genital human alpha papillomaviruses by PCR using novel broad-spectrum  
516 GP5+ and GP6+ primers. *J Clin Microbiol* **46**:1050-9.
- 517 18. **Schuurman, R., D. Brambilla, T. de Groot, D. Huang, S. Land, J. Bremer, I. Benders,**  
518 **and C. A. Boucher.** 2002. Underestimation of HIV type 1 drug resistance mutations:  
519 results from the ENVA-2 genotyping proficiency program. *AIDS Res Hum Retroviruses*  
520 **18**:243-8.
- 521 19. **Smith, J. S., L. Lindsay, B. Hoots, J. Keys, S. Franceschi, R. Winer, and G. M.**  
522 **Clifford.** 2007. Human papillomavirus type distribution in invasive cervical cancer and  
523 high-grade cervical lesions: a meta-analysis update. *Int J Cancer* **121**:621-32.
- 524 20. **Snijders, P. J., C. J. Hogewoning, A. T. Hesselink, J. Berkhof, F. J. Voorhorst, M. C.**  
525 **Bleeker, and C. J. Meijer.** 2006. Determination of viral load thresholds in cervical  
526 scrapings to rule out CIN 3 in HPV 16, 18, 31 and 33-positive women with normal  
527 cytology. *Int J Cancer* **119**:1102-7.
- 528 21. **Soderlund-Strand, A., J. Carlson, and J. Dillner.** 2009. Modified general primer PCR  
529 system for sensitive detection of multiple types of oncogenic human papillomavirus. *J Clin*  
530 *Microbiol* **47**:541-6.
- 531 22. **Sotlar, K., D. Diemer, A. Dethleffs, Y. Hack, A. Stubner, N. Vollmer, S. Menton, M.**  
532 **Menton, K. Dietz, D. Wallwiener, R. Kandolf, and B. Bultmann.** 2004. Detection and  
533 typing of human papillomavirus by e6 nested multiplex PCR. *J Clin Microbiol* **42**:3176-84.
- 534 23. **Stevens, M. P., S. M. Garland, and S. N. Tabrizi.** 2008. Development and validation of a  
535 real-time PCR assay specifically detecting human papillomavirus 52 using the Roche  
536 LightCycler 480 system. *J Virol Methods* **147**:290-6.
- 537 24. **Valentine-Thon, E., A. M. van Loon, J. Schirm, J. Reid, P. E. Klapper, and G. M.**  
538 **Cleator.** 2001. European proficiency testing program for molecular detection and  
539 quantitation of hepatitis B virus DNA. *J Clin Microbiol* **39**:4407-12.
- 540 25. **van den Brule, A. J., R. Pol, N. Fransen-Daalmeijer, L. M. Schouls, C. J. Meijer, and**  
541 **P. J. Snijders.** 2002. GP5+/6+ PCR followed by reverse line blot analysis enables rapid  
542 and high-throughput identification of human papillomavirus genotypes. *J Clin Microbiol*  
543 **40**:779-87.
- 544 26. **WHO** 2005, posting date. WHO technical workshop on the role of laboratory detection of  
545 human papillomavirus in global disease prevention and control. World Health Organisation.  
546 [Online.]  
547  
548

549  
550  
551  
552

553 **TABLE 1. HPV DNA proficiency panel composition and HPV testing results**

| HPV types                    | HPV genome equivalents per 5 µl | Percent correct data sets*<br>(N)                 |
|------------------------------|---------------------------------|---|
| 16                           | 500                             | <b>94,9 (74 / 78)</b>                             |
| 16                           | 50                              | 89,7 (70 / 78)                                    |
| 16                           | 5                               | 67,9 (53 / 78)                                    |
| 18                           | 500                             | <b>92,2 (71 / 77)</b>                             |
| 18                           | 50                              | 92,2 (71 / 77)                                    |
| 18                           | 5                               | 59,7 (46 / 77)                                    |
| 6                            | 500                             | <b>91,3 (63 / 69)</b>                             |
| 6                            | 50                              | 81,1 (56 / 69)                                    |
| 11                           | 500                             | <b>88,4 (61 / 69)</b>                             |
| 11                           | 50                              | 94,2 (65 / 69)                                    |
| 31                           | 500                             | <b>86,4 (64 / 74)</b>                             |
| 31                           | 50                              | 67,6 (50 / 74)                                    |
| 33                           | 500                             | <b>90,5 (67 / 74)</b>                             |
| 33                           | 50                              | 86,5 (64 / 74)                                    |
| 35                           | 500                             | <b>86,5 (64 / 74)</b>                             |
| 35                           | 50                              | 78,4 (58 / 74)                                    |
| 39                           | 500                             | <b>90,5 (38 / 42)</b>                             |
| 39                           | 50                              | 69,0 (29 / 42)                                    |
| 45                           | 500                             | <b>89,0 (65 / 73)</b>                             |
| 45                           | 50                              | 80,8 (59 / 73)                                    |
| 51                           | 500                             | <b>88,9 (64 / 72)</b>                             |
| 51                           | 50                              | 75 (54 / 72)                                      |
| 52                           | 500                             | <b>85,1 (63 / 74)</b>                             |
| 52                           | 50                              | 78,4 (58 / 74)                                    |
| 56                           | 500                             | <b>75,3 (55 / 73)</b>                             |
| 56                           | 50                              | 68,5 (50 / 73)                                    |
| 58                           | 500                             | <b>90,5 (67 / 74)</b>                             |
| 58                           | 50                              | 75,7 (56 / 74)                                    |
| 59                           | 500                             | <b>72,6 (53 / 73)</b>                             |
| 59                           | 50                              | 65,7 (48 / 73)                                    |
| 66                           | 500                             | <b>84,6 (55 / 65)</b>                             |
| 66                           | 50                              | 77,3 (51 / 65)                                    |
| 68                           | 500                             | <b>37,9 (11 / 29)</b>                             |
| 68                           | 50                              | 34,4 (10 / 29)                                    |
| 16, 45, 52, 33               | 500                             | <b>58,9 (46 / 78)<sup>a</sup></b>                 |
| 16, 45, 52, 33               | 50                              | 47,4 (37 / 78) <sup>a</sup>                       |
| 11, 18, 31, 51               | 500                             | <b>72,7 (56 / 77)<sup>a</sup></b>                 |
| 11, 18, 31, 51               | 50                              | 59,7 (46 / 77) <sup>a</sup>                       |
| 35, 39, 59, 66               | 500                             | <b>59,7 (44 / 74)<sup>b</sup></b>                 |
| 35, 39, 59, 66               | 50                              | 50 (37 / 74) <sup>b</sup>                         |
| 6, 56, 58, 68                | 500                             | <b>50,0 (37 / 74)<sup>b</sup></b>                 |
| 6, 56, 58, 68                | 50                              | 41,9 (31 / 74) <sup>b</sup>                       |
| None                         | 0                               | <b>92,5 (74 / 80)</b>                             |
| HPV 16 Cervical cancer cells | 2000 cells                      | 34,3 (23 / 67)<br>(3 false positives)             |
| HPV-negative cells           | 0                               | 65,7 (44 / 67)<br>(6 false positives, 17 invalid) |

554 \* Data sets detecting correct type as claimed, no false positive type detected.

555 <sup>a</sup> Including data sets generated by type specific HPV 16 / HPV 18 PCR.

556 <sup>b</sup> Data sets known not to detect HPV 39 or HPV 68 are considered as correct when the other HPV types in the  
557 sample are detected.

558 c) The plasmid concentration that is equivalent to 50 genome copies (IU) varied from 0,53fg to 0,67fg/5 µl because of  
559 small variation in the length of the HPV genome and use of different cloning vectors. HPV68 had only an L1 plasmid  
560 and the plasmid concentration equivalent to 50 genome copies was therefore 0,23fg/5µl. The background concentration  
561 of human DNA was in all samples 50ng/5 µl.

562  
563

**TABLE 2. Proficiency of detecting HPV types tested for, by laboratory**

| HPV assay type               | Number of data sets | HPV region targeted (primers) | No. of proficient data sets |                    |                    |                  |                |
|------------------------------|---------------------|-------------------------------|-----------------------------|--------------------|--------------------|------------------|----------------|
|                              |                     |                               | 100% proficient             | 99-90 % proficient | 89-80 % proficient | <80 % proficient | Not proficient |
| All assays                   | 73                  | L1/E1/E6/E7                   | 19                          | 10                 | 5                  | 11               | 28             |
| Linear Array (Roche)         | 15                  | L1 (PGMY)                     | 6                           | 5                  | 0                  | 0                | 4              |
| CLART HPV 2 (Genomica)       | 6                   | L1 (PGMY)                     | 0                           | 0                  | 2                  | 0                | 4              |
| InnoLiPA (Innogenetics)      | 6                   | L1 (SPF10)                    | 0                           | 1                  | 1                  | 0                | 4              |
| PGMY-LINEBLOT                | 6                   | L1 (PGMY)                     | 1                           | 1                  | 0                  | 2                | 2              |
| In-house Type-specific PCR   | 6                   | L1 / E6 / E7                  | 2                           | 0                  | 0                  | 1                | 3              |
| DNA chip (Biocore)           | 4                   | L1                            | 0                           | 0                  | 0                  | 3                | 1              |
| In-house Lineblot (Snijders) | 4                   | L1 (GP)                       | 0                           | 1                  | 0                  | 2                | 1              |
| In house PCR Luminex         | 4                   | L1 (GP)                       | 3                           | 0                  | 0                  | 0                | 1              |
| In house PCR Luminex         | 4                   | E6 / E7                       | 2                           | 0                  | 0                  | 0                | 2              |
| In-house Microarray          | 3                   | L1 / E7                       | 1                           | 0                  | 0                  | 0                | 2              |
| PCR-RFLP                     | 3                   | L1                            | 0                           | 0                  | 0                  | 2                | 1              |
| Microarray (Genetel)         | 2                   | L1                            | 2                           | 0                  | 0                  | 0                | 0              |
| DEIA LiPA assays (DDL)       | 2                   | L1 (SPF 10)                   | 0                           | 0                  | 0                  | 0                | 2              |
| In house PCR EIA             | 2                   | L1                            | 0                           | 0                  | 1                  | 0                | 1              |
| Papillocheck Microarray      | 1                   | E1                            | 1                           | 0                  | 0                  | 0                | 0              |
| Type specific PCR (GenoID)   | 1                   | L1                            | 0                           | 1                  | 0                  | 0                | 0              |
| In-house PCR Luminex         | 1                   | L1 (PGMY-GP)                  | 0                           | 0                  | 0                  | 1                | 0              |
| PCR Luminex (Multimetrix)    | 1                   | L1 (GP)                       | 0                           | 0                  | 1                  | 0                | 0              |
| PCR EIA (GenoID)             | 1                   | L1                            | 0                           | 1                  | 0                  | 0                | 0              |
| In-house PCR sequencing      | 1                   | L1 (PGMY-GP)                  | 1                           | 0                  | 0                  | 0                | 0              |

564  
565  
566  
567  
568  
569  
570

Table restricted to assays testing for more than 12 types.

571 **TABLE 3 False positive HPV types detected, by assay**

572

| HPV assay type               | Number of data sets | HPV region targeted (primers) | No. of false positive samples per data set |          |           |           |             |
|------------------------------|---------------------|-------------------------------|--|----------|-----------|-----------|-------------|
|                              |                     |                               | 0 samples                                  | 1 sample | 2 samples | 3 samples | > 3 samples |
| All assays                   | 80                  | L1/E1/E6/E7                   | 34   | 16       | 9         | 9         | 12          |
| Linear Array (Roche)         | 15                  | L1 (PGMY)                     | 6  | 5        | 2         | 2         | 0           |
| CLART HPV 2 (Genomica)       | 6                   | L1 (PGMY)                     | 1  | 1        | 1         | 0         | 3           |
| InnoLiPA (Innogenetics)      | 6                   | L1 (SPF10)                    | 1  | 1        | 0         | 2         | 2           |
| PGMY-LINEBLOT                | 6                   | L1 (PGMY)                     | 3  | 0        | 0         | 3         | 0           |
| In-house Type-specific PCR   | 7                   | L1 / E6 / E7                  | 1  | 3        | 1         | 0         | 2           |
| In-house 16 /18 specific PCR | 6                   | E6 / E7                       | 5  | 0        | 1         | 0         | 0           |
| DNA chip (Biocore)           | 4                   | L1                            | 1  | 2        | 0         | 1         | 0           |
| In-house Lineblot (Snijders) | 4                   | L1 (GP)                       | 2  | 1        | 0         | 0         | 1           |
| In house PCR Luminex         | 4                   | L1 (GP)                       | 3  | 0        | 0         | 0         | 1           |
| In house PCR Luminex         | 4                   | E6 / E7                       | 2  | 0        | 1         | 0         | 1           |
| In-house Microarray          | 3                   | L1 / E7                       | 0  | 1        | 1         | 0         | 1           |
| PCR-RFLP                     | 3                   | L1                            | 1  | 1        | 1         | 0         | 0           |
| Microarray (Genetel)         | 2                   | L1                            | 2  | 0        | 0         | 0         | 0           |
| DEIA LiPA assays (DDL)       | 2                   | L1 (SPF 10)                   | 0  | 0        | 1         | 1         | 0           |
| In house PCR EIA             | 2                   | L1                            | 0  | 1        | 0         | 0         | 1           |
| Papillocheck Microarray      | 1                   | E1                            | 1  | 0        | 0         | 0         | 0           |
| Type specific PCR (GenoID)   | 1                   | L1                            | 1  | 0        | 0         | 0         | 0           |
| In-house PCR Luminex         | 1                   | L1 (PGMY-GP)                  | 1  | 0        | 0         | 0         | 0           |
| PCR Luminex (Multimetrix)    | 1                   | L1 (GP)                       | 1  | 0        | 0         | 0         | 0           |
| PCR EIA (GenoID)             | 1                   | L1                            | 1  | 0        | 0         | 0         | 0           |
| In-house PCR sequencing      | 1                   | L1 (PGMY-GP)                  | 1  | 0        | 0         | 0         | 0           |

573

574 <sup>a</sup>Data including the 2 extraction control samples, that were not included in the proficiency

575 evaluation

576

577

578

579

580

581

582

583

584

585 **Table 4a:** HPV IU/GE detected per 5 µl in both single and multiple infections (commercial assays)

| HPV type | HPV IU /GE | All Assay (%) | Linear Array* (Roche) | CLART HPV 2 (Genomica) | InnoLiPA (Innogenetics) | DNA chip (biocore) | Microarray (Genetel) | Papillocheck Microarray | Luminex (Multimetrix) | PCR-EIA (GenoID) | Type specific PCR (GenoID) |
|----------|------------|---------------|-----------------------|------------------------|-------------------------|--------------------|----------------------|-------------------------|-----------------------|------------------|----------------------------|
| 16       | 5          | 50 / 79 (63)  | 7 / 15                | 4/6                    | 2 / 6                   | 4 / 4              | 2 / 2                |                         |                       |                  | 1 / 1                      |
| 16       | 50         | 69 / 79 (87)  | 15 / 15               | 5/6                    | 3 / 6                   |                    |                      | 1 / 1                   | 1 / 1                 |                  |                            |
| 16       | 500        | 78 / 79 (99)  |                       |                        | 6 / 6                   |                    |                      |                         |                       | 1 / 1            |                            |
| 18       | 5          | 41 / 78 (53)  | 4 / 15                | 1/6                    | 5 / 6                   | 4 / 4              | 2 / 2                |                         | 1 / 1                 |                  | 1 / 1                      |
| 18       | 50         | 69 / 78 (88)  | 14 / 15               | 6/6                    |                         |                    |                      |                         |                       | 1 / 1            |                            |
| 18       | 500        | 75 / 78 (96)  | 15 / 15               |                        |                         |                    |                      | nt                      |                       |                  |                            |
| 6        | 50         | 48 / 70 (69)  | 6 / 15                | 6 / 6                  | 6 / 6                   | 4 / 4              | 2 / 2                | 1 / 1                   |                       |                  | 1 / 1                      |
| 6        | 500        | 62 / 70 (88)  | 15 / 15               |                        |                         |                    |                      |                         |                       | 1 / 1            |                            |
| 11       | 50         | 56 / 70 (80)  | 6 / 15                | 6 / 6                  | 5 / 6                   | 4 / 4              | 2 / 2                | 1 / 1                   | 1 / 1                 | 1 / 1            | 1 / 1                      |
| 11       | 500        | 67 / 70 (96)  | 14 / 15               |                        |                         |                    |                      |                         |                       |                  |                            |
| 31       | 50         | 36 / 75 (48)  | 6 / 15                | 3 / 6                  | 5 / 6                   |                    | 2 / 2                | 1 / 1                   |                       | 1 / 1            | 1 / 1                      |
| 31       | 500        | 61 / 75 (81)  | 15 / 15               | 4 / 6                  |                         |                    |                      |                         |                       |                  |                            |
| 33       | 50         | 55 / 75 (73)  | 7 / 15                | 5 / 6                  | 6 / 6                   | 4 / 4              | 2 / 2                |                         | 1 / 1                 | 1 / 1            | 1 / 1                      |
| 33       | 500        | 70 / 75 (93)  | 15 / 15               |                        |                         |                    |                      | 1 / 1                   |                       |                  |                            |
| 35       | 50         | 50 / 75 (67)  | 7 / 15                | 4 / 6                  | 5 / 6                   | 4 / 4              | 1 / 2                |                         | 1 / 1                 | 1 / 1            | 1 / 1                      |
| 35       | 500        | 65 / 75 (87)  | 14 / 15               |                        | 6 / 6                   |                    | 2 / 2                | 1 / 1                   |                       |                  |                            |
| 39       | 50         | 25 / 42 (60)  |                       |                        | 5 / 6                   | 1 / 4              | 1 / 2                | 1 / 1                   | 1 / 1                 |                  | 1 / 1                      |
| 39       | 500        | 38 / 42 (90)  | Nt                    | nt                     | 6 / 6                   | 3 / 4              | 2 / 2                |                         |                       | 1 / 1            |                            |
| 45       | 50         | 48 / 74 (65)  | 7 / 15                | 1 / 6                  | 2 / 6                   | 4 / 4              |                      | 1 / 1                   | 1 / 1                 | 1 / 1            | 1 / 1                      |
| 45       | 500        | 63 / 74 (85)  | 15 / 15               | 2 / 6                  | 5 / 6                   |                    | 2 / 2                |                         |                       |                  |                            |
| 51       | 50         | 49 / 73 (67)  | 7 / 15                | 6 / 6                  | 5 / 6                   | 2 / 4              | 2 / 2                | 1 / 1                   | 1 / 1                 | 1 / 1            | 1 / 1                      |
| 51       | 500        | 64 / 73 (88)  | 15 / 15               |                        |                         |                    |                      |                         |                       |                  |                            |
| 52       | 50         | 40 / 75 (53)  | 3 / 15                | 4 / 6                  | 1 / 6                   | 2 / 4              | 2 / 2                |                         | 1 / 1                 |                  | 1 / 1                      |
| 52       | 500        | 50 / 75 (67)  | 9 / 15                |                        |                         | 3 / 4              |                      | 1 / 1                   |                       | 1 / 1            |                            |
| 56       | 50         | 41 / 74 (55)  | 4 / 15                | 1 / 6                  | 6 / 6                   |                    | 2 / 2                | 1 / 1                   | 1 / 1                 | 1 / 1            | 1 / 1                      |
| 56       | 500        | 56 / 74 (76)  | 14 / 15               | 2 / 6                  |                         |                    |                      |                         |                       |                  |                            |
| 58       | 50         | 48 / 75 (64)  | 7 / 15                | 5 / 6                  | 1 / 6                   | 3 / 4              | 2 / 2                |                         |                       | 1 / 1            | 1 / 1                      |
| 58       | 500        | 68 / 75 (91)  | 15 / 15               | 6 / 6                  | 4 / 6                   | 4 / 4              |                      | 1 / 1                   | 1 / 1                 |                  |                            |
| 59       | 50         | 42 / 74 (57)  | 7 / 15                | 4 / 6                  | 1 / 6                   |                    | 2 / 2                | 1 / 1                   |                       | 1 / 1            | 1 / 1                      |
| 59       | 500        | 55 / 74 (74)  | 15 / 15               |                        |                         |                    |                      |                         | 1 / 1                 |                  |                            |
| 66       | 50         | 44 / 66 (67)  | 6 / 15                | 6 / 6                  | 6 / 6                   |                    | 1 / 2                | 1 / 1                   | 1 / 1                 | 1 / 1            | 1 / 1                      |
| 66       | 500        | 58 / 66 (88)  | 14 / 15               |                        |                         |                    | 2 / 2                |                         |                       |                  |                            |
| 68       | 50         | 7 / 29 (24)   |                       |                        |                         |                    | 1 / 2                |                         |                       |                  |                            |
| 68       | 500        | 10 / 29 (34)  | Nt                    | nt                     | 1 / 5                   |                    | 2 / 2                | nt                      |                       |                  |                            |

586  
587

588  
589

**Table 4b:** HPV IU or GE detected per 5 µl in, both single and multiple infections (in-house assays).

| HPV type | HPV IU /GE | PGMY - CHUV | Type specific PCR | Lineblot | Luminex (GP) | Luminex (E6/E7) | Microarray | PCR-RFLP | DEIA LIPA | PCR-EIA | Luminex (PGMY-GP) | PCR sequencing |
|----------|------------|-------------|-------------------|----------|--------------|-----------------|------------|----------|-----------|---------|-------------------|----------------|
| 16       | 5          | 4/7         | 6/7               | 2/4      | 3/4          | 3/4             | 2/3        | 1/3      | 2/2       | 1/2     | 1/1               | 1/1            |
| 16       | 50         | 6/7         |                   | 3/4      | 4/4          |                 | 3/3        | 2/3      |           | 2/2     |                   |                |
| 16       | 500        | 7/7         | 7/7               | 4/4      |              | 4/4             |            | 3/3      |           |         |                   |                |
| 18       | 5          | 2/7         | 5/7               | 1/4      | 4/4          | 2/4             | 2/3        |          | 2/2       |         | 1/1               | 1/1            |
| 18       | 50         | 6/7         |                   | 4/4      |              | 4/4             | 3/3        |          |           | 1/2     |                   |                |
| 18       | 500        |             | 7/7               |          |              |                 |            | 2/3      |           | 2/2     |                   |                |
| 6        | 50         | 4/7         | 3/5               | 1/4      | 4/4          | 1/2             | 1/2        | 3/3      | 2/2       | 1/2     | 1/1               | 1/1            |
| 6        | 500        | 5/7         | 5/5               |          |              | 2/2             |            |          |           |         |                   |                |
| 11       | 50         | 5/7         | 4/5               | 4/4      | 4/4          | 1/2             | 2/2        | 3/3      | 2/2       | 2/2     | 1/1               | 1/1            |
| 11       | 500        | 7/7         | 5/5               |          |              |                 |            |          |           |         |                   |                |
| 31       | 50         |             | 5/7               | 2/4      | 1/4          | 2/4             | 1/3        | 1/3      | 2/2       | 2/2     |                   | 1/1            |
| 31       | 500        | 5/7         | 6/7               | 4/4      | 4/4          | 3/4             | 3/3        | 2/3      |           |         |                   |                |
| 33       | 50         | 2/7         | 7/7               | 3/4      | 4/4          | 3/4             | 2/3        | 2/3      | 2/2       | 2/2     |                   | 1/1            |
| 33       | 500        | 5/7         |                   | 4/4      |              |                 | 3/3        | 3/3      |           |         |                   |                |
| 35       | 50         | 2/7         | 6/7               | 4/4      | 4/4          | 2/4             | 2/3        |          | 2/2       | 2/2     | 1/1               | 1/1            |
| 35       | 500        | 5/7         |                   |          |              | 4/4             |            |          |           |         |                   |                |
| 39       | 50         |             | 4/6               |          | 4/4          | 3/4             | 2/3        |          | 2/2       |         |                   |                |
| 39       | 500        | nt          | 6/6               | 3/4      |              | 4/4             |            |          |           | 1/1     | nt                | nt             |
| 45       | 50         | 6/7         | 5/7               | 4/4      | 4/4          | 4/4             | 2/3        |          | 2/2       | 1/1     | 1/1               | 1/1            |
| 45       | 500        |             |                   |          |              |                 | 3/3        |          |           |         |                   |                |
| 51       | 50         | 5/7         | 5/6               | 1/4      | 4/4          | 2/4             | 1/3        | 1/3      | 2/2       |         | 1/1               | 1/1            |
| 51       | 500        | 7/7         | 6/6               | 3/4      |              | 3/4             |            |          |           | 1/1     |                   |                |
| 52       | 50         | 5/7         | 7/7               |          | 4/4          | 4/4             | 3/3        |          | 2/2       |         |                   | 1/1            |
| 52       | 500        |             |                   | 1/4      |              |                 |            |          |           |         |                   |                |
| 56       | 50         | 1/7         | 5/6               | 4/4      | 4/4          | 4/4             | 1/3        |          | 2/2       | 1/2     | 1/1               | 1/1            |
| 56       | 500        | 4/7         |                   |          |              |                 |            |          |           | 2/2     |                   |                |
| 58       | 50         | 5/7         | 7/7               | 3/4      | 3/4          | 2/4             | 2/3        | 2/3      | 2/2       |         | 1/1               | 1/1            |
| 58       | 500        |             |                   | 4/4      | 4/4          | 3/4             |            |          |           | 2/2     |                   |                |
| 59       | 50         | 6/7         | 5/7               | 3/4      | 3/4          | 2/4             | 2/3        |          | 1/2       | 1/1     | 1/1               | 1/1            |
| 59       | 500        | 7/7         |                   |          | 4/4          | 3/4             |            |          | 2/2       |         |                   |                |
| 66       | 50         | 5/7         |                   | 2/4      | 4/4          | 2/3             | 2/3        | 1/3      | 2/2       | 1/1     | 1/1               | 1/1            |
| 66       | 500        | 6/7         | nt                | 4/4      |              | 3/3             | 3/3        |          |           |         |                   |                |
| 68       | 50         |             |                   |          | 4/4          |                 | 1/2        |          | 1/2       |         |                   |                |
| 68       | 500        |             |                   |          |              | nt              |            | nt       | 2/2       |         | nt                | nt             |



590 Includes laboratories with multiple false positives. Detection with input volume 50  $\mu$ l classified as  
591 data for the next 10-fold dilution compared to input with 5  $\mu$ l. Input with 10 or 15  $\mu$ l classified as  
592 same dilution compared to input with 5  $\mu$ l.  
593 \* 8 laboratories used 50  $\mu$ l input volume in Linear Array.  
594 One InnoLiPA assay does not detect HPV 68.  
595  
596