LUND UNIVERSITY

Radiations from a Water Jet Plasma Source

Fullagar, Wilfred; Uhlig, Jens; Walczak, Monika; Canton, Sophie; Wahlström, Claes-Göran; Sundström, Villy

2007

Document Version:
Other version

Link to publication

Citation for published version (APA):
Fullagar, W., Uhlig, J., Walczak, M., Canton, S., Wahlström, C.-G., & Sundström, V. (2007). *Radiations from a Water Jet Plasma Source*. 7.

Total number of authors:
6

General rights
Unless other specific re-use rights are stated the following general rights apply:
Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.
• Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
• You may not further distribute the material or use it for any profit-making activity or commercial gain
• You may freely distribute the URL identifying the publication in the public portal

Read more about Creative commons licenses: https://creativecommons.org/licenses/

Take down policy
If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

LUND UNIVERSITY

PO Box 117
221 00 Lund
+46 46-222 00 00

Radiations from a Water Jet Plasma Source

Wilfred Fullagar, Jens Uhlig, Monika Walczak, Sophie Canton, Claes-Göran Wahlström, Villy Sundström

Division of Chemical Physics
Division of Atomic Physics



X-ray shadow image of a leaf



X-ray shadow image of a fly head

Laser system

low power branch of LLC terawatt facility (June 2007)



Output beam:
800 nm, ≥~35 fs, ≤~300 mJ/pulse, 10 Hz, ~4 cm ϕ

Typical temporal contrast (courtesy Filip Lindau)

X-rays



Manual scan of jet through focus

Goals:

In-house subpicosecond chemical structure dynamics, via
- pump-probe EXAFS
- pump-probe Laue crystallography



Shadow of ightbulb filament & 250µm Cu grid
Blurring ⇒ source ~30 µm



A typical setup, shield chamber removed

Source ambitions:	This source:
* Everyday local development access	**Yes** (LLC collaboration)
* Simplicity → in-house development & maintenance	**Yes**
* Broadband radiation	**Yes** (Maxwellian, T = 10-50 x10^6 K)
* Sub-picosecond burst	**Expected** (believed limited by (jet ϕ)/(e⁻ velocity), or laser pulse)
* No emission lines (→ detection & mechanism issues)	**Yes** (oxygen Kα too soft for typical filter transmission)
* Adequate X-ray flux	**Yes (EXAFS) but requires detection development**
* jitter-free laser synchronisation	**Yes**
* Collimation	**No (needed for Laue xtlgrphy; see electron beams, below)**
* Small source size	**Yes** (multi-shot images ⇒ ~30 µm in target plane)



Compressor scan: X-ray yield & autocorrelator FWHM overlaid



A typical mono-temperature spectrum & fit (table at right)



Instrumental resolution on X-ray streak camera



Single shot Ti K_α XAS edge spectrum (5 µm Ti foil)

Adaptable to typical kHz laser

Handy bunch of numbers (corresp. 10 Hz , 1 atm He operation)

Energy/pulse (mJ)	optical pulse duration	N_0 (ph/[eV.sr.sec])	T (K)
40	~200 fs (optimum)	71500	17.4x10^6
1.6	~200 fs (optimum)	14702	12.7x10^6
40	~40 fs (shortest)	29200	11.5x10^6

Selected relevant literature:
* R.J. Tompkins et al, Rev.Sci. Inst 69(9), 3113 (1998)
* K. Eidmann *et al, Europhys Lett.,* 55(3) 334 (2001)

Add s-pol prepulse, several 10s ps early, 1:1 pwr ratio
⇒ X-ray-free, controllable preplasma (vast parameter space)



Electron beams away from and through the water jet

Also! : dramatic change of interaction geometry for continuum X-rays
(typical literature ⇒ 30° - 60° incidence for assorted K_α)

Clear flashes on GdOS screen after ~1cm air
⇒ observe strong scatter by 50 µm Be window



Collimated:
including strong filter scatter & assoc 2ndary radiation <10° FWHM



Generates 5.9keV Kα emission line through 5 mm Ti X-ray attentuation foil.

Electrons



No prepulse
X-ray pinhole shadow.
(magnet present)

With prepulse
B-field pointing up
(NB lots scatter, vert rescale)

With prepulse
B-field pointing down
(NB lots scatter, vert rescale)

Deflects like electron in magnetic field (detector view from source)



Prepulse delay scan through t=0. I→r ≡~2ns.
⇒ e-beam requires p-polarised pulse in preplasma.



Electron ranges (CSDA, ESTAR, NIST). ρ^{-1} (cm³/g):
H_2O = 1; He = 5590, Be = 0.541, air = 769, Si = 0.429
Suggests observed beams are 20 – 200 keV.



Chirp scan, no prepulse
No electrons
X-rays only, as above



Repeat, with prepulse
e⁻ beams at shortest pulses
Note rescale to include saturation

Strong e⁻ generation when s- & p- pulses briefest. Suggests:
* proton motion during p-pulse unimportant
* e⁻ beam emission = non-resonant phenomenon
* e- beam pulse duration ← energy range, distance from src

Selected literature precedents:
* S. Bastiani *et al*, *Phys. Rev. E*, 56(6), 7179 (1997)
* X.-Y. Peng *et al*, *Chin. Phys. Lett.*, 21(4), 693 (2004)

Possible future?



Selected relevant literature:
* A. Rousse et al, Phys. Rev. Lett., 93(13), 135005 (2004)
* W.P. Leemans *et al*, *Nature Physics*, October(2), 696 (2006)

Protons??

H⁺ emission conceivable with long pulses, high NA focus
For example:



H⁺ jets moving to left, PIC simulation
F. Brunel, Phys. Fluids, 31(9) 2716 (1988)



Simulated electron spectra
with & without mobile H⁺
P. Gibbon, PRL 73(5), 664 (1994)



Circuit observes all ionization in chamber → strong signal when hitting jet.
But! : only partial correlation with X-ray yield.... ***why?***

Protons stopped in chamber by He (→ ionisation) : see below
However : 524 eV oxygen K_α (1/e in He → 12 cm), & possible other contibutions

Proton range in He (p-STAR, NIST)
ρ^{-1} = 5990 cm³/g
⇒ H⁺ < ~1 MeV stopped in chamber